

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052322\
 Data File : VY009004.D
 Acq On : 23 May 2022 14:21
 Operator : KP/MD
 Sample : N2868-03
 Misc : 6.48g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P017-SS001-1218-01

Quant Time: May 24 01:47:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	118325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	212702	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	206900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	88881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	78369	64.354	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.700%	
35) Dibromofluoromethane	7.716	113	68174	51.543	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.080%	
50) Toluene-d8	10.179	98	268349	54.995	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.980%	
62) 4-Bromofluorobenzene	12.477	95	78805	40.036	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 80.080%	
Target Compounds						
16) Acetone	3.948	43	3958	15.614	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052322\
Data File : VY009004.D
Acq On : 23 May 2022 14:21
Operator : KP/MD
Sample : N2868-03
Misc : 6.48g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P017-SS001-1218-01

Quant Time: May 24 01:47:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

