

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052322\
 Data File : VY009036.D
 Acq On : 24 May 2022 03:34
 Operator : KP/MD
 Sample : N2948-19
 Misc : 2.80g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P036-SS001-1824-01

Quant Time: May 24 04:08:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	113552	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	206232	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	191097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	81656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	70495	60.321	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.640%	
35) Dibromofluoromethane	7.716	113	65471	51.052	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.100%	
50) Toluene-d8	10.179	98	254624	53.819	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.640%	
62) 4-Bromofluorobenzene	12.483	95	72194	37.828	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 75.660%	
Target Compounds						Qvalue
16) Acetone	3.948	43	10409	42.789	ug/l	99
25) 2-Butanone	6.990	43	3171	7.298	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052322\
Data File : VY009036.D
Acq On : 24 May 2022 03:34
Operator : KP/MD
Sample : N2948-19
Misc : 2.80g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P036-SS001-1824-01

Quant Time: May 24 04:08:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

