

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052322\
 Data File : VY009040.D
 Acq On : 24 May 2022 05:08
 Operator : KP/MD
 Sample : N2948-23
 Misc : 5.28g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P037-SS001-0612-01

Quant Time: May 24 07:48:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	83616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	145838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	130323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	53962	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	49796	57.865	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	115.720%	
35) Dibromofluoromethane	7.716	113	45785	50.486	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	100.980%	
50) Toluene-d8	10.179	98	175402	52.427	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	104.860%	
62) 4-Bromofluorobenzene	12.483	95	48507	35.942	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	71.880%	
Target Compounds						
					Qvalue	
70) Styrene	12.044	104	3672	1.051	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052322\
Data File : VY009040.D
Acq On : 24 May 2022 05:08
Operator : KP/MD
Sample : N2948-23
Misc : 5.28g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P037-SS001-0612-01

Quant Time: May 24 07:48:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

