

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052322\
 Data File : VY009042.D
 Acq On : 24 May 2022 05:55
 Operator : KP/MD
 Sample : N2948-28
 Misc : 5.61g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P037-SS001-1824-01

Quant Time: May 24 07:49:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	109411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	202102	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	191805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	83454	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	73136	64.950	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.900%
35) Dibromofluoromethane	7.716	113	65065	51.772	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.540%
50) Toluene-d8	10.179	98	252881	54.543	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.080%
62) 4-Bromofluorobenzene	12.483	95	73253	39.167	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052322\
Data File : VY009042.D
Acq On : 24 May 2022 05:55
Operator : KP/MD
Sample : N2948-28
Misc : 5.61g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P037-SS001-1824-01

Quant Time: May 24 07:49:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

