

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
 Data File : VY009052.D
 Acq On : 24 May 2022 11:26
 Operator : KP/MD
 Sample : N2948-14
 Misc : 5.38g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P036-SS001-0612-01

Quant Time: May 25 02:50:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	118780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	216165	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	204729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	84750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	75138	61.464	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.920%
35) Dibromofluoromethane	7.715	113	69338	51.583	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.160%
50) Toluene-d8	10.178	98	269003	54.246	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.500%
62) 4-Bromofluorobenzene	12.483	95	76446	38.215	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.440%
Target Compounds						
16) Acetone	3.948	43	6663	26.184	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
Data File : VY009052.D
Acq On : 24 May 2022 11:26
Operator : KP/MD
Sample : N2948-14
Misc : 5.38g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P036-SS001-0612-01

Quant Time: May 25 02:50:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

