

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
 Data File : VY009069.D
 Acq On : 24 May 2022 18:03
 Operator : KP/MD
 Sample : N2948-20
 Misc : 4.29g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P036-SS001-2430-01

Quant Time: May 25 02:55:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	104440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	193771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	182862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	77935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	67617	62.907	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 125.820%			
35) Dibromofluoromethane	7.716	113	62549	51.910	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 103.820%			
50) Toluene-d8	10.179	98	238350	53.619	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 107.240%			
62) 4-Bromofluorobenzene	12.483	95	68988	38.473	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 76.940%			
Target Compounds						
16) Acetone	3.954	43	4056	18.128	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
Data File : VY009069.D
Acq On : 24 May 2022 18:03
Operator : KP/MD
Sample : N2948-20
Misc : 4.29g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P036-SS001-2430-01

Quant Time: May 25 02:55:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

