

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
 Data File : VY009073.D
 Acq On : 24 May 2022 19:37
 Operator : KP/MD
 Sample : N2981-06
 Misc : 2.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-14(3.5-4)

Quant Time: May 25 02:56:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	106877	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	191703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	177978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	77317	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	66828	60.755	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.500%
35) Dibromofluoromethane	7.716	113	60587	50.824	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.640%
50) Toluene-d8	10.179	98	237026	53.897	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.800%
62) 4-Bromofluorobenzene	12.477	95	67438	38.014	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
Data File : VY009073.D
Acq On : 24 May 2022 19:37
Operator : KP/MD
Sample : N2981-06
Misc : 2.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-14(3.5-4)

Quant Time: May 25 02:56:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

