

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
 Data File : VY009076.D
 Acq On : 24 May 2022 20:47
 Operator : KP/MD
 Sample : N2981-09
 Misc : 2.45g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-11(0.5-1)

Quant Time: May 25 02:57:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	104593	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	189743	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	174329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	71516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	66468	61.747	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.500%	
35) Dibromofluoromethane	7.716	113	60801	51.531	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.060%	
50) Toluene-d8	10.179	98	232537	53.422	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.840%	
62) 4-Bromofluorobenzene	12.483	95	64289	36.613	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 73.220%	
Target Compounds						
16) Acetone	3.948	43	4505	20.105	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
Data File : VY009076.D
Acq On : 24 May 2022 20:47
Operator : KP/MD
Sample : N2981-09
Misc : 2.45g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-11(0.5-1)

Quant Time: May 25 02:57:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

