

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
 Data File : VY009049.D
 Acq On : 24 May 2022 09:50
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 25 02:48:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	71	0.00
2 T	Dichlorodifluoromethane	0.457	0.447	2.2	66	0.00
3 P	Chloromethane	0.655	0.719	-9.8	74	0.00
4 C	Vinyl Chloride	0.971	1.020	-5.0#	71	0.00
5 T	Bromomethane	0.951	0.821	13.7	71	0.00
6 T	Chloroethane	0.688	0.702	-2.0	73	0.00
7 T	Trichlorofluoromethane	0.824	0.819	0.6	69	0.00
8 T	Diethyl Ether	0.258	0.253	1.9	70	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.507	0.508	-0.2	69	0.00
10 T	Methyl Iodide	0.523	0.548	-4.8	63	0.00
11 T	Tert butyl alcohol	0.052	0.042	19.2	58	0.00
12 CM	1,1-Dichloroethene	0.470	0.470	0.0	69	0.00
13 T	Acrolein	0.017	0.007	58.8#	53	0.01
14 T	Allyl chloride	0.709	0.655	7.6	63	0.00
15 T	Acrylonitrile	0.144	0.134	6.9	66	0.00
16 T	Acetone	0.107	0.121	-13.1	82	0.00
17 T	Carbon Disulfide	1.536	1.519	1.1	67	0.00
18 T	Methyl Acetate	0.334	0.300	10.2	63	0.00
19 T	Methyl tert-butyl Ether	1.412	1.319	6.6	65	0.00
20 T	Methylene Chloride	0.788	0.625	20.7	71	0.00
21 T	trans-1,2-Dichloroethene	0.604	0.570	5.6	67	0.00
22 T	Diisopropyl ether	1.703	1.651	3.1	67	0.00
23 T	Vinyl Acetate	1.126	1.097	2.6	65	0.00
24 P	1,1-Dichloroethane	1.018	0.979	3.8	67	0.00
25 T	2-Butanone	0.191	0.191	0.0	68	0.00
26 T	2,2-Dichloropropane	0.938	0.948	-1.1	71	0.00
27 T	cis-1,2-Dichloroethene	0.680	0.646	5.0	65	0.00
28 T	Bromochloromethane	0.350	0.344	1.7	71	0.00
29 T	Tetrahydrofuran	0.130	0.123	5.4	64	0.00
30 C	Chloroform	1.099	1.041	5.3#	66	0.00
31 T	Cyclohexane	1.007	0.952	5.5	67	0.00
32 T	1,1,1-Trichloroethane	0.997	0.971	2.6	66	0.00
33 S	1,2-Dichloroethane-d4	0.515	0.467	9.3	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	70	0.00
35 S	Dibromofluoromethane	0.311	0.295	5.1	79	0.00
36 T	1,1-Dichloropropene	0.529	0.516	2.5	66	0.00
37 T	Ethyl Acetate	0.274	0.248	9.5	62	0.00
38 T	Carbon Tetrachloride	0.554	0.532	4.0	65	0.00
39 T	Methylcyclohexane	0.683	0.666	2.5	64	0.00
40 TM	Benzene	1.526	1.472	3.5	65	0.00
41 T	Methacrylonitrile	0.142	0.133	6.3	62	0.00
42 TM	1,2-Dichloroethane	0.448	0.443	1.1	67	0.00
43 T	Isopropyl Acetate	0.510	0.500	2.0	65	0.00
44 TM	Trichloroethene	0.404	0.379	6.2	64	0.00
45 C	1,2-Dichloropropane	0.366	0.366	0.0	66	0.00
46 T	Dibromomethane	0.233	0.222	4.7	64	0.00
47 T	Bromodichloromethane	0.532	0.517	2.8	65	0.00
48 T	Methyl methacrylate	0.228	0.233	-2.2	65	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
 Data File : VY009049.D
 Acq On : 24 May 2022 09:50
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 25 02:48:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	58	0.00
50 S	Toluene-d8	1.147	0.985	14.1	78	0.00
51 T	4-Methyl-2-Pentanone	0.282	0.280	0.7	64	0.00
52 CM	Toluene	1.066	1.040	2.4#	63	0.00
53 T	t-1,3-Dichloropropene	0.565	0.570	-0.9	65	0.00
54 T	cis-1,3-Dichloropropene	0.608	0.603	0.8	66	0.00
55 T	1,1,2-Trichloroethane	0.326	0.308	5.5	62	0.00
56 T	Ethyl methacrylate	0.417	0.413	1.0	62	0.00
57 T	1,3-Dichloropropane	0.532	0.523	1.7	65	0.00
58 T	2-Chloroethyl Vinyl ether	0.180	0.185	-2.8	66	0.00
59 T	2-Hexanone	0.193	0.203	-5.2	66	0.00
60 T	Dibromochloromethane	0.395	0.381	3.5	63	0.00
61 T	1,2-Dibromoethane	0.323	0.306	5.3	62	0.00
62 S	4-Bromofluorobenzene	0.463	0.497	-7.3	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	68	0.00
64 T	Tetrachloroethene	0.364	0.341	6.3	63	0.00
65 PM	Chlorobenzene	1.148	1.085	5.5	62	0.00
66 T	1,1,1,2-Tetrachloroethane	0.420	0.407	3.1	62	0.00
67 C	Ethyl Benzene	2.068	2.073	-0.2#	63	0.00
68 T	m/p-Xylenes	0.840	0.870	-3.6	63	0.00
69 T	o-Xylene	0.780	0.801	-2.7	63	0.00
70 T	Styrene	1.341	1.422	-6.0	62	0.00
71 P	Bromoform	0.279	0.275	1.4	61	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
73 T	Isopropylbenzene	3.739	3.377	9.7	62	0.00
74 T	N-amyl acetate	0.863	0.840	2.7	66	0.00
75 P	1,1,2,2-Tetrachloroethane	0.770	0.650	15.6	61	0.00
76 T	1,2,3-Trichloropropane	0.537	0.492	8.4	62	0.00
77 T	Bromobenzene	0.867	0.740	14.6	61	0.00
78 T	n-propylbenzene	4.320	4.282	0.9	101	0.00
79 T	2-Chlorotoluene	2.492	2.283	8.4	62	0.00
80 T	1,3,5-Trimethylbenzene	3.134	2.960	5.6	62	0.00
81 T	trans-1,4-Dichloro-2-butene	0.262	0.237	9.5	63	0.00
82 T	4-Chlorotoluene	2.665	2.560	3.9	62	0.00
83 T	tert-Butylbenzene	2.780	2.591	6.8	62	0.00
84 T	1,2,4-Trimethylbenzene	3.111	3.010	3.2	63	0.00
85 T	sec-Butylbenzene	4.254	4.137	2.8	63	0.00
86 T	p-Isopropyltoluene	3.547	3.505	1.2	63	0.00
87 T	1,3-Dichlorobenzene	1.849	1.737	6.1	62	0.00
88 T	1,4-Dichlorobenzene	1.857	1.744	6.1	63	0.00
89 T	n-Butylbenzene	3.245	3.245	0.0	66	0.00
90 T	Hexachloroethane	0.665	0.619	6.9	63	0.00
91 T	1,2-Dichlorobenzene	1.632	1.510	7.5	63	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.101	15.8	59	0.00
93 T	1,2,4-Trichlorobenzene	0.896	0.779	13.1	62	0.00
94 T	Hexachlorobutadiene	0.488	0.423	13.3	62	0.00
95 T	Naphthalene	1.799	1.546	14.1	59	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
Data File : VY009049.D
Acq On : 24 May 2022 09:50
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 25 02:48:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.778	0.674	13.4	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 4