

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
 Data File : VY009075.D
 Acq On : 24 May 2022 20:23
 Operator : KP/MD
 Sample : N2981-08
 Misc : 2.13g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-12(2.5-3)

Quant Time: May 25 02:57:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	105235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	188855	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	174042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	72557	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	66544	61.441	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.880%
35) Dibromofluoromethane	7.716	113	61139	52.061	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.120%
50) Toluene-d8	10.179	98	230895	53.294	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.580%
62) 4-Bromofluorobenzene	12.483	95	65731	37.611	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052422\
Data File : VY009075.D
Acq On : 24 May 2022 20:23
Operator : KP/MD
Sample : N2981-08
Misc : 2.13g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-12(2.5-3)

Quant Time: May 25 02:57:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY009075.D
Acq: 24 May 2022 20:23




Instrument :
MSVOA_Y
ClientSampleId :
SB-12(2.5-3)


Tgt Ion:168 Resp: 105235
Ion Ratio Lower Upper
168 100
99 55.8 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 61.441 ug/l
RT: 8.137 min Scan# 1036
Delta R.T. -0.006 min
Lab File: VY009075.D
Acq: 24 May 2022 20:23

Tgt Ion: 65 Resp: 66544
Ion Ratio Lower Upper
65 100
67 53.4 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY009075.D

Acq: 24 May 2022 20:23

Instrument :

MSVOA_Y

ClientSampleId :

SB-12(2.5-3)





Tgt Ion:114 Resp: 188855

Ion Ratio Lower Upper

114 100

63 18.1 0.0 43.2

88 13.9 0.0 31.8



#35

Dibromofluoromethane

Concen: 52.061 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY009075.D

Acq: 24 May 2022 20:23




Tgt Ion:113 Resp: 61139


Ion Ratio Lower Upper

113 100

111 101.2 82.6 124.0

192 18.6 17.6 26.4









#50
Toluene-d8
Concen: 53.294 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY009075.D
Acq: 24 May 2022 20:23




Instrument :
MSVOA_Y
ClientSampleId :
SB-12(2.5-3)


Tgt Ion: 98 Resp: 230895
Ion Ratio Lower Upper
98 100
100 64.3 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 37.611 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY009075.D
Acq: 24 May 2022 20:23

Tgt Ion: 95 Resp: 65731
Ion Ratio Lower Upper
95 100
174 89.3 0.0 177.0
176 87.5 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 117
Delta R.T. -0.000 min
Lab File: VY009075.D
Acq: 24 May 2022 20:23

Instrument :
MSVOA_Y
ClientSampleId :
SB-12(2.5-3)

Tgt Ion:117 Resp: 174042
Ion Ratio Lower Upper
117 100
82 51.9 43.0 64.4
119 31.4 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY009075.D
Acq: 24 May 2022 20:23

Tgt Ion:152 Resp: 72557
Ion Ratio Lower Upper
152 100
115 56.1 28.2 84.7
150 159.7 0.0 347.6

