

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052424\
 Data File : VY018367.D
 Acq On : 24 May 2024 15:56
 Operator : SY/MD
 Sample : P2330-22
 Misc : 4.92g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 H-3(5.0-6.0)

Quant Time: May 25 03:25:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 23 01:40:36 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	112702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	219866	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	173918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	45415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	60574	56.730	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.460%
35) Dibromofluoromethane	7.704	113	65214	50.863	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.720%
50) Toluene-d8	10.173	98	248054	49.622	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.240%
62) 4-Bromofluorobenzene	12.477	95	61221	39.635	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.280%
Target Compounds						
					Qvalue	
16) Acetone	3.936	43	2649	12.062	ug/l	97
17) Carbon Disulfide	4.180	76	16287	4.891	ug/l	98
20) Methylene Chloride	4.692	84	13635	6.579	ug/l	88
84) 1,2,4-Trimethylbenzene	13.117	105	3796	1.376	ug/l	99
95) Naphthalene	15.239	128	74846	50.023	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052424\
Data File : VY018367.D
Acq On : 24 May 2024 15:56
Operator : SY/MD
Sample : P2330-22
Misc : 4.92g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
H-3(5.0-6.0)

Quant Time: May 25 03:25:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
Quant Title : SW846 8260
QLast Update : Thu May 23 01:40:36 2024
Response via : Initial Calibration

