

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052424\
 Data File : VY018372.D
 Acq On : 24 May 2024 18:16
 Operator : SY/MD
 Sample : P2615-03
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RBR200035

Quant Time: May 25 03:27:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 23 01:40:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	45944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	90634	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	72872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	24572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	25727	59.104	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.200%
35) Dibromofluoromethane	7.710	113	26785	50.678	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.360%
50) Toluene-d8	10.173	98	100860	48.946	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.900%
62) 4-Bromofluorobenzene	12.477	95	27540	43.253	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.500%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	7978	11.776	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052424\
Data File : VY018372.D
Acq On : 24 May 2024 18:16
Operator : SY/MD
Sample : P2615-03
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RBR200035

Quant Time: May 25 03:27:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
Quant Title : SW846 8260
QLast Update : Thu May 23 01:40:36 2024
Response via : Initial Calibration

