

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052522\
 Data File : VY009083.D
 Acq On : 25 May 2022 12:35
 Operator : KP/MD
 Sample : N2981-11
 Misc : 2.71g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10(0.5-1)

Quant Time: May 26 04:27:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	113407	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	204847	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	188162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	76337	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	69750	59.760	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.520%
35) Dibromofluoromethane	7.716	113	64949	50.987	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.980%
50) Toluene-d8	10.179	98	254496	54.156	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.320%
62) 4-Bromofluorobenzene	12.483	95	69358	36.588	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	73.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052522\
Data File : VY009083.D
Acq On : 25 May 2022 12:35
Operator : KP/MD
Sample : N2981-11
Misc : 2.71g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10(0.5-1)

Quant Time: May 26 04:27:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

