

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052522\
 Data File : VY009087.D
 Acq On : 25 May 2022 14:09
 Operator : KP/MD
 Sample : N2981-15
 Misc : 2.47g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-8(0.5-1)

Quant Time: May 26 04:28:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	106130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	194193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	183096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	78305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	66843	61.196	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.400%
35) Dibromofluoromethane	7.716	113	62448	51.714	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.420%
50) Toluene-d8	10.179	98	242263	54.381	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.760%
62) 4-Bromofluorobenzene	12.477	95	67801	37.729	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.460%
Target Compounds						
16) Acetone	3.942	43	2474	10.881	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052522\
Data File : VY009087.D
Acq On : 25 May 2022 14:09
Operator : KP/MD
Sample : N2981-15
Misc : 2.47g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-8(0.5-1)

Quant Time: May 26 04:28:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

