

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052522\
 Data File : VY009080.D
 Acq On : 25 May 2022 11:01
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 26 04:26:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	68	0.00
2 T	Dichlorodifluoromethane	0.457	0.504	-10.3	72	0.00
3 P	Chloromethane	0.655	0.766	-16.9	76	0.00
4 C	Vinyl Chloride	0.971	1.105	-13.8#	75	0.00
5 T	Bromomethane	0.951	0.875	8.0	73	0.00
6 T	Chloroethane	0.688	0.751	-9.2	75	0.00
7 T	Trichlorofluoromethane	0.824	0.876	-6.3	71	0.00
8 T	Diethyl Ether	0.258	0.263	-1.9	70	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.507	0.550	-8.5	72	0.00
10 T	Methyl Iodide	0.523	0.603	-15.3	67	0.00
11 T	Tert butyl alcohol	0.052	0.048	7.7	65	0.00
12 CM	1,1-Dichloroethene	0.470	0.501	-6.6#	72	0.01
13 T	Acrolein	0.017	0.008	52.9#	58	0.00
14 T	Allyl chloride	0.709	0.752	-6.1	70	0.00
15 T	Acrylonitrile	0.144	0.152	-5.6	72	0.00
16 T	Acetone	0.107	0.119	-11.2	77	0.00
17 T	Carbon Disulfide	1.536	1.625	-5.8	70	0.00
18 T	Methyl Acetate	0.334	0.348	-4.2	70	0.00
19 T	Methyl tert-butyl Ether	1.412	1.473	-4.3	70	0.00
20 T	Methylene Chloride	0.788	0.606	23.1	67	0.00
21 T	trans-1,2-Dichloroethene	0.604	0.623	-3.1	71	0.00
22 T	Diisopropyl ether	1.703	1.856	-9.0	72	0.00
23 T	Vinyl Acetate	1.126	1.233	-9.5	70	0.00
24 P	1,1-Dichloroethane	1.018	1.085	-6.6	72	0.00
25 T	2-Butanone	0.191	0.207	-8.4	71	0.00
26 T	2,2-Dichloropropane	0.938	1.040	-10.9	75	0.00
27 T	cis-1,2-Dichloroethene	0.680	0.713	-4.9	69	0.00
28 T	Bromochloromethane	0.350	0.356	-1.7	71	0.00
29 T	Tetrahydrofuran	0.130	0.139	-6.9	70	0.00
30 C	Chloroform	1.099	1.147	-4.4#	70	0.00
31 T	Cyclohexane	1.007	1.021	-1.4	69	0.00
32 T	1,1,1-Trichloroethane	0.997	1.064	-6.7	70	0.00
33 S	1,2-Dichloroethane-d4	0.515	0.436	15.3	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	67	0.00
35 S	Dibromofluoromethane	0.311	0.278	10.6	71	0.00
36 T	1,1-Dichloropropene	0.529	0.574	-8.5	70	0.00
37 T	Ethyl Acetate	0.274	0.285	-4.0	68	0.00
38 T	Carbon Tetrachloride	0.554	0.589	-6.3	69	0.00
39 T	Methylcyclohexane	0.683	0.736	-7.8	67	0.00
40 TM	Benzene	1.526	1.633	-7.0	69	0.00
41 T	Methacrylonitrile	0.142	0.126	11.3	55	0.00
42 TM	1,2-Dichloroethane	0.448	0.491	-9.6	71	0.00
43 T	Isopropyl Acetate	0.510	0.566	-11.0	70	0.00
44 TM	Trichloroethene	0.404	0.422	-4.5	68	0.00
45 C	1,2-Dichloropropane	0.366	0.399	-9.0#	69	0.00
46 T	Dibromomethane	0.233	0.248	-6.4	68	0.00
47 T	Bromodichloromethane	0.532	0.576	-8.3	70	0.00
48 T	Methyl methacrylate	0.228	0.257	-12.7	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052522\
 Data File : VY009080.D
 Acq On : 25 May 2022 11:01
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 26 04:26:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	68	0.00
50 S	Toluene-d8	1.147	0.950	17.2	72	0.00
51 T	4-Methyl-2-Pentanone	0.282	0.320	-13.5	69	0.00
52 CM	Toluene	1.066	1.170	-9.8#	67	0.00
53 T	t-1,3-Dichloropropene	0.565	0.632	-11.9	69	0.00
54 T	cis-1,3-Dichloropropene	0.608	0.666	-9.5	70	0.00
55 T	1,1,2-Trichloroethane	0.326	0.349	-7.1	67	0.00
56 T	Ethyl methacrylate	0.417	0.457	-9.6	66	0.00
57 T	1,3-Dichloropropane	0.532	0.585	-10.0	69	0.00
58 T	2-Chloroethyl Vinyl ether	0.180	0.192	-6.7	66	0.00
59 T	2-Hexanone	0.193	0.226	-17.1	69	0.00
60 T	Dibromochloromethane	0.395	0.428	-8.4	67	0.00
61 T	1,2-Dibromoethane	0.323	0.348	-7.7	67	0.00
62 S	4-Bromofluorobenzene	0.463	0.482	-4.1	66	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	65	0.00
64 T	Tetrachloroethene	0.364	0.378	-3.8	66	0.00
65 PM	Chlorobenzene	1.148	1.231	-7.2	67	0.00
66 T	1,1,1,2-Tetrachloroethane	0.420	0.465	-10.7	68	0.00
67 C	Ethyl Benzene	2.068	2.330	-12.7#	67	0.00
68 T	m/p-Xylenes	0.840	0.981	-16.8	67	0.00
69 T	o-Xylene	0.780	0.909	-16.5	67	0.00
70 T	Styrene	1.341	1.606	-19.8	66	0.00
71 P	Bromoform	0.279	0.310	-11.1	65	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
73 T	Isopropylbenzene	3.739	3.679	1.6	67	0.00
74 T	N-amyl acetate	0.863	0.935	-8.3	73	0.00
75 P	1,1,2,2-Tetrachloroethane	0.770	0.723	6.1	67	0.00
76 T	1,2,3-Trichloropropane	0.537	0.540	-0.6	67	0.00
77 T	Bromobenzene	0.867	0.807	6.9	65	0.00
78 T	n-propylbenzene	4.320	4.642	-7.5	108	0.00
79 T	2-Chlorotoluene	2.492	2.476	0.6	66	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.242	-3.4	67	0.00
81 T	trans-1,4-Dichloro-2-butene	0.262	0.259	1.1	68	0.00
82 T	4-Chlorotoluene	2.665	2.791	-4.7	67	0.00
83 T	tert-Butylbenzene	2.780	2.820	-1.4	66	0.00
84 T	1,2,4-Trimethylbenzene	3.111	3.266	-5.0	67	0.00
85 T	sec-Butylbenzene	4.254	4.539	-6.7	68	0.00
86 T	p-Isopropyltoluene	3.547	3.839	-8.2	68	0.00
87 T	1,3-Dichlorobenzene	1.849	1.910	-3.3	68	0.00
88 T	1,4-Dichlorobenzene	1.857	1.911	-2.9	68	0.00
89 T	n-Butylbenzene	3.245	3.531	-8.8	70	0.00
90 T	Hexachloroethane	0.665	0.664	0.2	67	0.00
91 T	1,2-Dichlorobenzene	1.632	1.647	-0.9	67	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.108	10.0	62	0.00
93 T	1,2,4-Trichlorobenzene	0.896	0.845	5.7	66	0.00
94 T	Hexachlorobutadiene	0.488	0.456	6.6	66	0.00
95 T	Naphthalene	1.799	1.694	5.8	64	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052522\
Data File : VY009080.D
Acq On : 25 May 2022 11:01
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 26 04:26:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.778	0.729	6.3	66	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6