

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052523\
 Data File : VY013874.D
 Acq On : 25 May 2023 15:36
 Operator : KP/MD
 Sample : 02869-01
 Misc : 3.25g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP18

Quant Time: May 26 04:34:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	11130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	15870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	13738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	5108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	7261	56.390	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.780%
35) Dibromofluoromethane	7.722	113	5995	58.494	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.980%
50) Toluene-d8	10.179	98	18665	47.917	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.840%
62) 4-Bromofluorobenzene	12.483	95	5167	39.906	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.820%
Target Compounds						
20) Methylene Chloride	4.704	84	1849	11.700	ug/l	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052523\
Data File : VY013874.D
Acq On : 25 May 2023 15:36
Operator : KP/MD
Sample : 02869-01
Misc : 3.25g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP18

Quant Time: May 26 04:34:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

