

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052523\
 Data File : VY013878.D
 Acq On : 25 May 2023 17:10
 Operator : KP/MD
 Sample : 02908-02
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB10B

Quant Time: May 26 04:34:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	14487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	22024	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	20267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	7740	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	10591	63.192	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.380%
35) Dibromofluoromethane	7.722	113	8523	59.923	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.840%
50) Toluene-d8	10.179	98	26693	49.379	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.760%
62) 4-Bromofluorobenzene	12.477	95	6842	38.077	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	76.160%
Target Compounds						
16) Acetone	3.936	43	4783	96.153	ug/l	93
20) Methylene Chloride	4.704	84	3669	19.240	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052523\
Data File : VY013878.D
Acq On : 25 May 2023 17:10
Operator : KP/MD
Sample : 02908-02
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB10B

Quant Time: May 26 04:34:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

