

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052523\
 Data File : VY013881.D
 Acq On : 25 May 2023 18:20
 Operator : KP/MD
 Sample : 02908-07RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB13RE

Quant Time: May 26 04:35:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	19665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	27420	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	22166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	7593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	10467	46.008	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.020%
35) Dibromofluoromethane	7.722	113	9350	52.801	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.600%
50) Toluene-d8	10.179	98	30893	45.902	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.800%
62) 4-Bromofluorobenzene	12.483	95	7278	32.533	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	65.060%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.716	84	3797	14.032	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052523\
Data File : VY013881.D
Acq On : 25 May 2023 18:20
Operator : KP/MD
Sample : 02908-07RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB13RE

Quant Time: May 26 04:35:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

