

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052523\
 Data File : VY013885.D
 Acq On : 25 May 2023 19:54
 Operator : KP/MD
 Sample : 02941-02
 Misc : 6.59g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TILCON-CLEANFILL-001

Quant Time: May 26 04:35:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	172953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	301545	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	298752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	136294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	127169	63.556	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.120%
35) Dibromofluoromethane	7.722	113	104090	53.451	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.900%
50) Toluene-d8	10.185	98	370466	50.053	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.100%
62) 4-Bromofluorobenzene	12.483	95	119326	48.502	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.000%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.729	84	28832	11.750	ug/l	96
52) Toluene	10.246	92	52675	10.141	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052523\
Data File : VY013885.D
Acq On : 25 May 2023 19:54
Operator : KP/MD
Sample : 02941-02
Misc : 6.59g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TILCON-CLEANFILL-001

Quant Time: May 26 04:35:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

