

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052621\
 Data File : VY004886.D
 Acq On : 26 May 2021 15:04
 Operator : SY/MD
 Sample : M2521-01RE
 Misc : 2.54G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-01-052521RE

Quant Time: May 27 01:16:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 11 04:13:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.795	168	180859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	334576	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	323590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	150618	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	131648	61.15	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 122.30%			
35) Dibromofluoromethane	7.722	113	105261	57.47	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 114.94%			
50) Toluene-d8	10.179	98	404372	48.87	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 97.74%			
62) 4-Bromofluorobenzene	12.483	95	151081	54.10	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 108.20%			
Target Compounds						Qvalue
11) Tert butyl alcohol	4.972	59	388	1.79	ug/l #	71
16) Acetone	3.954	43	6443	13.32	ug/l	93
20) Methylene Chloride	4.722	84	6221	Below Cal	#	87
25) 2-Butanone	7.009	43	1175	1.37	ug/l #	51
31) Cyclohexane	7.789	56	4412	1.27	ug/l #	1
36) 1,1-Dichloropropene	7.795	75	15324	4.47	ug/l #	51
38) Carbon Tetrachloride	7.789	117	18434	5.98	ug/l #	16
51) 4-Methyl-2-Pentanone	10.081	43	3375	1.44	ug/l	90
52) Toluene	10.246	92	17119	2.80	ug/l	99
55) 1,1,2-Trichloroethane	10.618	97	3460	1.78	ug/l #	18
59) 2-Hexanone	10.801	43	14323	9.01	ug/l #	31
67) Ethyl Benzene	11.593	91	69991	5.66	ug/l	99
68) m/p-Xylenes	11.697	106	110007	23.66	ug/l	99
69) o-Xylene	12.026	106	61430	14.14	ug/l	99
73) Isopropylbenzene	12.331	105	42284	3.72	ug/l	99
74) N-amyl acetate	12.124	43	54745	13.63	ug/l #	69
75) 1,1,2,2-Tetrachloroethane	12.642	83	11287	4.98	ug/l #	26
76) 1,2,3-Trichloropropane	12.733	75	2562	1.60	ug/l #	100
78) n-propylbenzene	12.672	91	122904	8.86	ug/l	99
79) 2-Chlorotoluene	12.733	91	37207	4.89	ug/l #	42
80) 1,3,5-Trimethylbenzene	12.812	105	175524	18.60	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.483	75	84896	92.01	ug/l #	14
82) 4-Chlorotoluene	12.812	91	19794	2.51	ug/l #	50
84) 1,2,4-Trimethylbenzene	13.123	105	517716	54.81	ug/l	99
85) sec-Butylbenzene	13.251	105	82379	7.03	ug/l	82
86) p-Isopropyltoluene	13.367	119	55416	5.46	ug/l	99
89) n-Butylbenzene	13.696	91	114378	11.52	ug/l #	72
90) Hexachloroethane	13.971	117	14450	10.11	ug/l #	14
92) 1,2-Dibromo-3-Chloropr...	14.300	75	871	2.06	ug/l	68
95) Naphthalene	15.233	128	15780	2.55	ug/l #	88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052621\
Data File : VY004886.D
Acq On : 26 May 2021 15:04
Operator : SY/MD
Sample : M2521-01RE
Misc : 2.54G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Quant Time: May 27 01:16:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
Quant Title : SW846 8260
QLast Update : Tue May 11 04:13:44 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.795 min Scan# 980
Delta R.T. 0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE







Tgt Ion:168 Resp: 180859
Ion Ratio Lower Upper
168 100
99 59.3 45.6 68.4



#11
Tert butyl alcohol
Concen: 1.79 ug/l
RT: 4.972 min Scan# 517
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 59 Resp: 388
Ion Ratio Lower Upper
59 100
57 0.0 8.7 13.1#









#16
Acetone
Concen: 13.32 ug/l
RT: 3.954 min Scan# 350
Delta R.T. -0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion: 43 Resp: 6443
Ion Ratio Lower Upper
43 100
58 25.6 23.6 35.4



#20
Methylene Chloride
Concen: Below Cal
RT: 4.722 min Scan# 476
Delta R.T. 0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04


Tgt Ion: 84 Resp: 6221
Ion Ratio Lower Upper
84 100
49 121.2 113.8 170.8
51 31.9 33.5 50.3#
86 66.2 51.1 76.7








#25
2-Butanone
Concen: 1.37 ug/l
RT: 7.009 min Scan# 851
Delta R.T. 0.012 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04


Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE





Tgt Ion: 43 Resp: 1175
Ion Ratio Lower Upper
43 100
72 0.0 19.4 29.0#




#31
Cyclohexane
Concen: 1.27 ug/l
RT: 7.789 min Scan# 979
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04



Tgt Ion: 56 Resp: 4412
Ion Ratio Lower Upper
56 100
69 162.8 24.4 36.6#
84 130.8 64.8 97.2#









#33
1,2-Dichloroethane-d4
Concen: 61.15 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion: 65 Resp: 131648
Ion Ratio Lower Upper
65 100
67 52.9 0.0 108.0



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.691 min Scan# 1127
Delta R.T. -0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04


Tgt Ion: 114 Resp: 334576
Ion Ratio Lower Upper
114 100
63 21.4 0.0 42.8
88 15.5 0.0 31.0








#35
Dibromofluoromethane
Concen: 57.47 ug/l
RT: 7.722 min Scan# 968
Delta R.T. -0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04


Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE





Tgt Ion:113 Resp: 105261
Ion Ratio Lower Upper
113 100
111 102.2 80.5 120.7
192 18.8 15.0 22.6




#36
1,1-Dichloropropene
Concen: 4.47 ug/l
RT: 7.795 min Scan# 980
Delta R.T. -0.128 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Tgt Ion: 75 Resp: 15324
Ion Ratio Lower Upper
75 100
110 9.8 17.0 50.9#
77 0.0 25.0 37.4#








#38
Carbon Tetrachloride
Concen: 5.98 ug/l
RT: 7.789 min Scan# 979
Delta R.T. -0.116 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE







Tgt Ion: 117 Resp: 18434
Ion Ratio Lower Upper
117 100
119 4.6 76.3 114.5#
121 0.0 23.0 34.6#



#50
Toluene-d8
Concen: 48.87 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 98 Resp: 404372
Ion Ratio Lower Upper
98 100
100 64.3 51.5 77.3









#51
4-Methyl-2-Pentanone
Concen: 1.44 ug/l
RT: 10.081 min Scan# 1355
Delta R.T. 0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion: 43 Resp: 3375
Ion Ratio Lower Upper
43 100
58 30.1 28.7 43.1



#52
Toluene
Concen: 2.80 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 92 Resp: 17119
Ion Ratio Lower Upper
92 100
91 171.9 138.2 207.4









#55
1,1,2-Trichloroethane
Concen: 1.78 ug/l
RT: 10.618 min Scan# 1443
Delta R.T. -0.030 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion:	97	Resp:	3460
Ion	Ratio	Lower	Upper
97	100		
83	8.3	71.9	107.9#
85	2.3	45.8	68.6#
99	0.0	51.1	76.7#



#59
2-Hexanone
Concen: 9.01 ug/l
RT: 10.801 min Scan# 1473
Delta R.T. -0.037 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion:	43	Resp:	14323
Ion	Ratio	Lower	Upper
43	100		
58	2.7	24.9	74.8#









#62
4-Bromofluorobenzene
Concen: 54.10 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion: 95 Resp: 151081
Ion Ratio Lower Upper
95 100
174 80.0 0.0 164.2
176 78.9 0.0 159.0



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.489 min Scan# 1586
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 117 Resp: 323590
Ion Ratio Lower Upper
117 100
82 57.0 45.4 68.2
119 31.8 25.8 38.6









#67
Ethyl Benzene
Concen: 5.66 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion: 91 Resp: 69991
Ion Ratio Lower Upper
91 100
106 30.4 24.0 36.0



#68
m/p-Xylenes
Concen: 23.66 ug/l
RT: 11.697 min Scan# 1620
Delta R.T. -0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion:106 Resp: 110007
Ion Ratio Lower Upper
106 100
91 202.5 163.6 245.4









#69
o-Xylene
Concen: 14.14 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. -0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion:106 Resp: 61430
Ion Ratio Lower Upper
106 100
91 216.4 108.9 326.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion:152 Resp: 150618
Ion Ratio Lower Upper
152 100
115 64.3 29.4 88.3
150 154.9 0.0 349.8









#73
Isopropylbenzene
Concen: 3.72 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion:105 Resp: 42284
Ion Ratio Lower Upper
105 100
120 26.8 13.1 39.1



#74
N-ethyl acetate
Concen: 13.63 ug/l
RT: 12.124 min Scan# 1690
Delta R.T. -0.024 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 43 Resp: 54745
Ion Ratio Lower Upper
43 100
70 17.6 30.2 45.2#
55 15.8 17.8 26.8#
61 0.0 17.6 26.4#









#75
1,1,2,2-Tetrachloroethane
Concen: 4.98 ug/l
RT: 12.642 min Scan# 1775
Delta R.T. 0.055 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion: 83 Resp: 11287
Ion Ratio Lower Upper
83 100
131 1.5 5.0 14.9#
85 0.0 32.1 96.5#



#76
1,2,3-Trichloropropane
Concen: 1.60 ug/l
RT: 12.733 min Scan# 1790
Delta R.T. 0.098 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 75 Resp: 2562
Ion Ratio Lower Upper
75 100
77 1420.9 0.0 0.0#





#78

n-propylbenzene

Concen: 8.86 ug/l

RT: 12.672 min Scan# 1780

Delta R.T. -0.000 min

Lab File: VY004886.D





Acq: 26 May 2021 15:04

Tgt Ion: 91 Resp: 122904

Ion Ratio Lower Upper

91 100

120 22.7 11.1 33.3



#79

2-Chlorotoluene

Concen: 4.89 ug/l

RT: 12.733 min Scan# 1790

Delta R.T. -0.024 min

Lab File: VY004886.D




Acq: 26 May 2021 15:04


Tgt Ion: 91 Resp: 37207

Ion Ratio Lower Upper

91 100

126 0.6 16.8 50.3#









#80
1,3,5-Trimethylbenzene
Concen: 18.60 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion:105 Resp: 175524
Ion Ratio Lower Upper
105 100
120 48.4 24.0 72.0



#81
trans-1,4-Dichloro-2-butene
Concen: 92.01 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.104 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 75 Resp: 84896
Ion Ratio Lower Upper
75 100
53 0.6 73.9 110.9#
89 2.2 35.5 53.3#









#82
4-Chlorotoluene
Concen: 2.51 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. -0.043 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion: 91 Resp: 19794
Ion Ratio Lower Upper
91 100
126 4.9 16.8 50.3#



#84
1,2,4-Trimethylbenzene
Concen: 54.81 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 105 Resp: 517716
Ion Ratio Lower Upper
105 100
120 45.6 22.5 67.5





#85

sec-Butylbenzene

Concen: 7.03 ug/l

RT: 13.251 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VY004886.D





Acq: 26 May 2021 15:04

Tgt Ion:105 Resp: 82379

Ion Ratio Lower Upper

105 100
134 28.1 10.0 29.8

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE



#86

p-Isopropyltoluene

Concen: 5.46 ug/l

RT: 13.367 min Scan# 1894

Delta R.T. -0.006 min




Lab File: VY004886.D


Acq: 26 May 2021 15:04

Tgt Ion:119 Resp: 55416

Ion Ratio Lower Upper

119 100
134 25.7 12.9 38.7
91 23.3 12.0 36.1









#89
n-Butylbenzene
Concen: 11.52 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. -0.000 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04




Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE


Tgt Ion: 91 Resp: 114378
Ion Ratio Lower Upper
91 100
92 39.8 26.4 79.0
134 0.0 12.6 37.6#



#90
Hexachloroethane
Concen: 10.11 ug/l
RT: 13.971 min Scan# 1993
Delta R.T. 0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 117 Resp: 14450
Ion Ratio Lower Upper
117 100
201 0.0 35.9 107.8#









#92
1,2-Dibromo-3-Chloropropane
Concen: 2.06 ug/l
RT: 14.300 min Scan# 2047
Delta R.T. -0.055 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-052521RE

Tgt Ion: 75 Resp: 871
Ion Ratio Lower Upper
75 100
155 104.0 39.0 117.0
157 65.9 50.2 150.7



#95
Naphthalene
Concen: 2.55 ug/l
RT: 15.233 min Scan# 2200
Delta R.T. -0.006 min
Lab File: VY004886.D
Acq: 26 May 2021 15:04

Tgt Ion: 128 Resp: 15780
Ion Ratio Lower Upper
128 100
127 16.7 10.2 15.2#
129 15.9 8.6 13.0#

