

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052622\
 Data File : VY009116.D
 Acq On : 26 May 2022 14:26
 Operator : KP/MD
 Sample : N2981-18
 Misc : 2.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-7(3.5-4)

Quant Time: May 27 03:24:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	102753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	180958	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	173719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	74130	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	62600	59.195	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.400%
35) Dibromofluoromethane	7.716	113	58318	51.826	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.660%
50) Toluene-d8	10.179	98	228080	54.942	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.880%
62) 4-Bromofluorobenzene	12.483	95	65281	38.983	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052622\
Data File : VY009116.D
Acq On : 26 May 2022 14:26
Operator : KP/MD
Sample : N2981-18
Misc : 2.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-7(3.5-4)

Quant Time: May 27 03:24:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

