

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052721\
 Data File : VY004900.D
 Acq On : 27 May 2021 15:00
 Operator : SY/MD
 Sample : M2535-09
 Misc : 5.62G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GHOST-5

Quant Time: May 27 15:06:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 11 04:13:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	177044	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	335868	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	308164	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	126791	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	122085	57.93	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.86%
35) Dibromofluoromethane	7.728	113	102206	55.59	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.18%
50) Toluene-d8	10.185	98	415002	49.96	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.92%
62) 4-Bromofluorobenzene	12.483	95	125494	44.76	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.52%
Target Compounds						
7) Trichlorofluoromethane	3.131	101	3866	1.21	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052721\
Data File : VY004900.D
Acq On : 27 May 2021 15:00
Operator : SY/MD
Sample : M2535-09
Misc : 5.62G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GHOST-5

Quant Time: May 27 15:06:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051021S.M
Quant Title : SW846 8260
QLast Update : Tue May 11 04:13:44 2021
Response via : Initial Calibration

