



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY052721\  
 Data File : VY004889.D  
 Acq On : 27 May 2021 10:00  
 Operator : SY/MD  
 Sample : BFB  
 Misc : 5.00G/5ML/MSVOA\_Y/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y051021S.M  
 Title : SW846 8260  
 Last Update : Tue May 11 04:13:44 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 21.6         | 25624      | PASS                |
| 75             | 95              | 30              | 60              | 54.9         | 65176      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 118613     | PASS                |
| 96             | 95              | 5               | 9               | 6.6          | 7780       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.2          | 1086       | PASS                |
| 174            | 95              | 50              | 100             | 79.3         | 94088      | PASS                |
| 175            | 174             | 5               | 9               | 7.3          | 6906       | PASS                |
| 176            | 174             | 95              | 101             | 94.9         | 89304      | PASS                |
| 177            | 176             | 5               | 9               | 7.1          | 6358       | PASS                |