

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv052820\
 Data File : VY002791.D
 Acq On : 28 May 2020 16:32
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 28 17:11:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 13:13:33 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.344	0.309	10.2	93	0.00
3 P	Chloromethane	0.381	0.365	4.2	100	0.00
4 C	Vinyl Chloride	0.402	0.396	1.5#	100	0.00
5 T	Bromomethane	0.301	0.298	1.0	101	0.00
6 T	Chloroethane	0.261	0.259	0.8	100	0.00
7 T	Trichlorofluoromethane	0.712	0.690	3.1	96	0.00
8 T	Diethyl Ether	0.278	0.267	4.0	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.423	0.430	-1.7	97	0.00
10 T	Methyl Iodide	0.592	0.586	1.0	91	0.00
11 T	Tert butyl alcohol	0.054	0.051	5.6	106	0.00
12 CM	1,1-Dichloroethene	0.425	0.431	-1.4#	99	0.00
13 T	Acrolein	0.051	0.048	5.9	95	0.00
14 T	Allyl chloride	0.610	0.652	-6.9	101	0.00
15 T	Acrylonitrile	0.125	0.135	-8.0	103	0.00
16 T	Acetone	0.102	0.104	-2.0	95	0.00
17 T	Carbon Disulfide	1.285	1.266	1.5	96	0.00
18 T	Methyl Acetate	0.276	0.279	-1.1	104	0.00
19 T	Methyl tert-butyl Ether	1.174	1.226	-4.4	101	0.00
20 T	Methylene Chloride	0.497	0.504	-1.4	103	0.00
21 T	trans-1,2-Dichloroethene	0.478	0.481	-0.6	98	0.00
22 T	Diisopropyl ether	1.268	1.360	-7.3	103	0.00
23 T	Vinyl Acetate	0.837	0.920	-9.9	102	0.00
24 P	1,1-Dichloroethane	0.742	0.778	-4.9	100	0.00
25 T	2-Butanone	0.159	0.168	-5.7	101	0.00
26 T	2,2-Dichloropropane	0.711	0.689	3.1	97	0.00
27 T	cis-1,2-Dichloroethene	0.536	0.549	-2.4	100	0.00
28 T	Bromochloromethane	0.304	0.319	-4.9	101	0.00
29 T	Tetrahydrofuran	0.103	0.112	-8.7	103	0.00
30 C	Chloroform	0.787	0.819	-4.1#	100	0.00
31 T	Cyclohexane	0.736	0.708	3.8	97	0.00
32 T	1,1,1-Trichloroethane	0.733	0.729	0.5	96	0.00
33 S	1,2-Dichloroethane-d4	0.434	0.451	-3.9	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.293	0.311	-6.1	101	0.00
36 T	1,1-Dichloropropene	0.421	0.421	0.0	99	0.00
37 T	Ethyl Acetate	0.229	0.245	-7.0	104	0.00
38 T	Carbon Tetrachloride	0.468	0.452	3.4	96	0.00
39 T	Methylcyclohexane	0.555	0.544	2.0	98	0.00
40 TM	Benzene	1.245	1.256	-0.9	99	0.00
41 T	Methacrylonitrile	0.123	0.157	-27.6#	130	0.00
42 TM	1,2-Dichloroethane	0.349	0.349	0.0	99	0.00
43 T	Isopropyl Acetate	0.434	0.459	-5.8	101	0.00
44 TM	Trichloroethene	0.393	0.392	0.3	98	0.00
45 C	1,2-Dichloropropane	0.298	0.308	-3.4#	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv052820\
 Data File : VY002791.D
 Acq On : 28 May 2020 16:32
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 28 17:11:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 13:13:33 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.177	0.183	-3.4	101	0.00
47 T	Bromodichloromethane	0.418	0.434	-3.8	100	0.00
48 T	Methyl methacrylate	0.196	0.209	-6.6	103	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	104	0.00
50 S	Toluene-d8	1.185	1.204	-1.6	99	0.00
51 T	4-Methyl-2-Pentanone	0.230	0.246	-7.0	103	0.00
52 CM	Toluene	0.814	0.810	0.5#	98	0.00
53 T	t-1,3-Dichloropropene	0.456	0.466	-2.2	99	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.530	-2.5	99	0.00
55 T	1,1,2-Trichloroethane	0.271	0.277	-2.2	99	0.00
56 T	Ethyl methacrylate	0.363	0.386	-6.3	102	0.00
57 T	1,3-Dichloropropane	0.446	0.458	-2.7	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.178	0.186	-4.5	104	0.00
59 T	2-Hexanone	0.163	0.173	-6.1	101	0.00
60 T	Dibromochloromethane	0.344	0.351	-2.0	99	0.00
61 T	1,2-Dibromoethane	0.268	0.272	-1.5	99	0.00
62 S	4-Bromofluorobenzene	0.405	0.419	-3.5	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.454	0.432	4.8	95	0.00
65 PM	Chlorobenzene	0.988	0.985	0.3	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.373	-0.3	99	0.00
67 C	Ethyl Benzene	1.713	1.686	1.6#	99	0.00
68 T	m/p-Xylenes	0.664	0.659	0.8	99	0.00
69 T	o-Xylene	0.627	0.620	1.1	98	0.00
70 T	Styrene	1.087	1.102	-1.4	100	0.00
71 P	Bromoform	0.249	0.249	0.0	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.204	3.163	1.3	99	0.00
74 T	N-amyl acetate	0.822	0.870	-5.8	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.562	0.585	-4.1	102	0.00
76 T	1,2,3-Trichloropropane	0.451	0.435	3.5	95	0.00
77 T	Bromobenzene	0.834	0.824	1.2	99	0.00
78 T	n-propylbenzene	3.701	3.670	0.8	99	0.00
79 T	2-Chlorotoluene	2.076	2.059	0.8	99	0.00
80 T	1,3,5-Trimethylbenzene	2.675	2.659	0.6	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.239	0.246	-2.9	100	0.00
82 T	4-Chlorotoluene	2.175	2.139	1.7	97	0.00
83 T	tert-Butylbenzene	2.386	2.379	0.3	99	0.00
84 T	1,2,4-Trimethylbenzene	2.690	2.679	0.4	98	0.00
85 T	sec-Butylbenzene	3.263	3.221	1.3	98	0.00
86 T	p-Isopropyltoluene	3.036	3.024	0.4	99	0.00
87 T	1,3-Dichlorobenzene	1.580	1.545	2.2	98	0.00
88 T	1,4-Dichlorobenzene	1.581	1.549	2.0	98	0.00
89 T	n-Butylbenzene	2.757	2.755	0.1	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv052820\
Data File : VY002791.D
Acq On : 28 May 2020 16:32
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 28 17:11:45 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
Quant Title : SW846 8260
QLast Update : Wed May 27 13:13:33 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.568	0.562	1.1	98	0.00
91 T	1,2-Dichlorobenzene	1.433	1.412	1.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.108	0.112	-3.7	102	0.00
93 T	1,2,4-Trichlorobenzene	1.080	1.071	0.8	99	0.00
94 T	Hexachlorobutadiene	0.613	0.604	1.5	98	0.00
95 T	Naphthalene	2.166	2.190	-1.1	100	0.00
96 T	1,2,3-Trichlorobenzene	0.965	0.955	1.0	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6