

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052824\
 Data File : VY018383.D
 Acq On : 28 May 2024 16:25
 Operator : SY/MD
 Sample : P2612-04
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT1486

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/29/2024
 Supervised By :Semsettin Yesilyurt 05/29/2024

Quant Time: May 29 01:20:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 23 01:40:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	1905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	3328	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	2030	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	998	55.296	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.600%	
35) Dibromofluoromethane	7.716	113	1114m	57.401	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.800%	
50) Toluene-d8	10.179	98	3806	50.301	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.600%	
62) 4-Bromofluorobenzene	12.483	95	447	19.119	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 38.240%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052824\
Data File : VY018383.D
Acq On : 28 May 2024 16:25
Operator : SY/MD
Sample : P2612-04
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT1486

Quant Time: May 29 01:20:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052224S.M
Quant Title : SW846 8260
QLast Update : Thu May 23 01:40:36 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/29/2024
Supervised By :Semsettin Yesilyurt 05/29/2024

