

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
 Data File : VY013920.D
 Acq On : 30 May 2023 15:32
 Operator : KP/MD
 Sample : 02852-06
 Misc : 5.49g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-32-G3-SO-33-051823

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 05/31/2023
 Supervised By :Mahesh Dadoda 05/31/2023

Quant Time: May 31 03:26:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	161591	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	275842	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	276421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	128390	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	118992	63.651	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 127.300%			
35) Dibromofluoromethane	7.716	113	94727	53.176	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 106.360%			
50) Toluene-d8	10.179	98	343031	50.665	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 101.340%			
62) 4-Bromofluorobenzene	12.483	95	111606	49.591	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 99.180%			
Target Compounds					Qvalue	
16) Acetone	3.960	43	27675	49.878	ug/l	99
20) Methylene Chloride	4.716	84	29610	13.181	ug/l	95
25) 2-Butanone	7.009	43	2932m	4.331	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
Data File : VY013920.D
Acq On : 30 May 2023 15:32
Operator : KP/MD
Sample : 02852-06
Misc : 5.49g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CB-32-G3-SO-33-051823

Quant Time: May 31 03:26:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/31/2023
Supervised By :Mahesh Dadoda 05/31/2023

