

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
 Data File : VY013926.D
 Acq On : 30 May 2023 17:52
 Operator : KP/MD
 Sample : 02852-12
 Misc : 6.88g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-32-G3-SO-55.5-051823-FD

Quant Time: May 31 03:27:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	154487	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.685	114	260777	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.483	117	254032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	116920	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	111170	62.201	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.400%	
35) Dibromofluoromethane	7.709	113	89149	52.935	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.880%	
50) Toluene-d8	10.178	98	324899	50.759	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.520%	
62) 4-Bromofluorobenzene	12.477	95	103773	48.774	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 97.540%	
Target Compounds						Qvalue
20) Methylene Chloride	4.704	84	16266	6.435	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
Data File : VY013926.D
Acq On : 30 May 2023 17:52
Operator : KP/MD
Sample : 02852-12
Misc : 6.88g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CB-32-G3-SO-55.5-051823-FD

Quant Time: May 31 03:27:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

