

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
 Data File : VY013927.D
 Acq On : 30 May 2023 18:16
 Operator : KP/MD
 Sample : 02852-13
 Misc : 5.73g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-32-G3-SO-62.5-051823

Quant Time: May 31 03:28:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	154863	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.685	114	266348	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.483	117	269747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	126070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	120060	67.012	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 134.020%	
35) Dibromofluoromethane	7.710	113	92522	53.789	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.580%	
50) Toluene-d8	10.172	98	329359	50.380	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.760%	
62) 4-Bromofluorobenzene	12.477	95	109602	50.436	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 100.880%	
Target Compounds						
						Qvalue
17) Carbon Disulfide	4.186	76	57022	12.611	ug/l	98
20) Methylene Chloride	4.704	84	16431	6.505	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
Data File : VY013927.D
Acq On : 30 May 2023 18:16
Operator : KP/MD
Sample : 02852-13
Misc : 5.73g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CB-32-G3-SO-62.5-051823

Quant Time: May 31 03:28:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

