

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
 Data File : VY013929.D
 Acq On : 30 May 2023 19:03
 Operator : KP/MD
 Sample : 02852-15
 Misc : 6.33g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-32-G3-SO-69.5-051823

Quant Time: May 31 03:28:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	157102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	257099	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.489	117	246295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	106000	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	99100	54.525	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.040%	
35) Dibromofluoromethane	7.715	113	81778	49.253	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.500%	
50) Toluene-d8	10.178	98	314704	49.870	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 99.740%	
62) 4-Bromofluorobenzene	12.483	95	96463	45.987	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 91.980%	
Target Compounds						Qvalue
17) Carbon Disulfide	4.198	76	45147	9.842	ug/l	97
20) Methylene Chloride	4.710	84	20478	8.604	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
Data File : VY013929.D
Acq On : 30 May 2023 19:03
Operator : KP/MD
Sample : 02852-15
Misc : 6.33g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CB-32-G3-SO-69.5-051823

Quant Time: May 31 03:28:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

