



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053023\
Data File : VY013909.D
Acq On : 30 May 2023 09:09
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Title : SW846 8260
Last Update : Wed May 24 05:25:47 2023



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1741

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.3	21814	PASS
75	95	30	60	53.6	57549	PASS
95	95	100	100	100.0	107368	PASS
96	95	5	9	6.4	6820	PASS
173	174	0.00	2	1.8	1757	PASS
174	95	50	100	93.5	100395	PASS
175	174	5	9	7.5	7569	PASS
176	174	95	101	97.0	97413	PASS
177	176	5	9	7.1	6944	PASS