

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053123\
 Data File : VY013939.D
 Acq On : 31 May 2023 16:35
 Operator : KP/MD
 Sample : 02955-11
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB21RE

Quant Time: May 31 17:33:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	5021	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	8360	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.483	117	6592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1369	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	5239	90.191	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	180.380%#
35) Dibromofluoromethane	7.710	113	3928	72.755	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	145.520%
50) Toluene-d8	10.179	98	10464	50.995	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.000%
62) 4-Bromofluorobenzene	12.477	95	2630	38.559	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053123\
Data File : VY013939.D
Acq On : 31 May 2023 16:35
Operator : KP/MD
Sample : 02955-11
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB21RE

Quant Time: May 31 17:33:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

