

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053123\
 Data File : VY013951.D
 Acq On : 31 May 2023 21:17
 Operator : KP/MD
 Sample : 02852-09RE
 Misc : 6.47g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CB-32-G3-SO-50.5-051823RE

Quant Time: Jun 01 03:32:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 24 05:25:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	19614	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	26399	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	20665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	7216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	10352	45.621	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.240%
35) Dibromofluoromethane	7.728	113	9065	53.171	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.340%
50) Toluene-d8	10.185	98	30183	46.581	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.160%
62) 4-Bromofluorobenzene	12.483	95	6892	31.999	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	64.000%
Target Compounds						
16) Acetone	3.954	43	1803	26.771	ug/l	91
44) Trichloroethene	8.947	130	493	2.454	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053123\
Data File : VY013951.D
Acq On : 31 May 2023 21:17
Operator : KP/MD
Sample : 02852-09RE
Misc : 6.47g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CB-32-G3-SO-50.5-051823RE

Quant Time: Jun 01 03:32:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Quant Title : SW846 8260
QLast Update : Wed May 24 05:25:47 2023
Response via : Initial Calibration

