



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY053123\
Data File : VY013933.D
Acq On : 31 May 2023 12:47
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052323S.M
Title : SW846 8260
Last Update : Wed May 24 05:25:47 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.6	21416	PASS
75	95	30	60	54.4	53915	PASS
95	95	100	100	100.0	99165	PASS
96	95	5	9	6.6	6579	PASS
173	174	0.00	2	1.2	1103	PASS
174	95	50	100	93.6	92861	PASS
175	174	5	9	7.4	6875	PASS
176	174	95	101	96.3	89421	PASS
177	176	5	9	6.7	5955	PASS