

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060123\
 Data File : VY013962.D
 Acq On : 01 Jun 2023 13:38
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jun 02 07:04:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	192669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	333037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	328829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	155698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	135560	53.151	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.300%	
35) Dibromofluoromethane	7.710	113	109507	51.740	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.480%	
50) Toluene-d8	10.179	98	404853	50.148	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.300%	
62) 4-Bromofluorobenzene	12.477	95	139874	50.821	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.640%	
Target Compounds						
					Qvalue	
95) Naphthalene	15.233	128	11660	5.361	ug/l	98
96) 1,2,3-Trichlorobenzene	15.428	180	4692	1.724	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060123\
Data File : VY013962.D
Acq On : 01 Jun 2023 13:38
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jun 02 07:04:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

