

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060123\
 Data File : VY013967.D
 Acq On : 01 Jun 2023 16:28
 Operator : KP/MD
 Sample : 02980-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-05302023

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/02/2023
 Supervised By :Mahesh Dadoda 06/02/2023

Quant Time: Jun 02 07:27:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	178964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	302877	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	300686	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	146309	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	129812	54.795	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.580%	
35) Dibromofluoromethane	7.709	113	101335	52.647	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.300%	
50) Toluene-d8	10.172	98	369607	50.341	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.680%	
62) 4-Bromofluorobenzene	12.477	95	133084	53.169	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 106.340%	
Target Compounds						
16) Acetone	3.948	43	3275m	4.041	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060123\
Data File : VY013967.D
Acq On : 01 Jun 2023 16:28
Operator : KP/MD
Sample : 02980-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-05302023

Quant Time: Jun 02 07:27:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/02/2023
Supervised By :Mahesh Dadoda 06/02/2023

