

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060123\
 Data File : VY013969.D
 Acq On : 01 Jun 2023 17:15
 Operator : KP/MD
 Sample : 02983-03RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB19RE

Quant Time: Jun 02 07:28:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	70203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	115584	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	116567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	53864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	53779	57.869	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.740%
35) Dibromofluoromethane	7.716	113	41477	56.466	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.940%
50) Toluene-d8	10.179	98	140802	50.253	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.500%
62) 4-Bromofluorobenzene	12.477	95	45642	47.782	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.560%
Target Compounds						
16) Acetone	3.942	43	8701	27.366	ug/l	98
20) Methylene Chloride	4.704	84	4879	1.825	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060123\
Data File : VY013969.D
Acq On : 01 Jun 2023 17:15
Operator : KP/MD
Sample : 02983-03RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB19RE

Quant Time: Jun 02 07:28:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

