


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv060220\
Data File : VY002810.D
Acq On : 02 Jun 2020 09:51
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
Title : SW846 8260
Last Update : Wed May 27 13:13:33 2020



AutoFind: Scans 1761, 1762, 1763; Background Corrected with Scan 1753

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	15585	PASS
75	95	30	60	48.2	47194	PASS
95	95	100	100	100.0	97861	PASS
96	95	5	9	6.7	6586	PASS
173	174	0.00	2	1.0	873	PASS
174	95	50	100	92.3	90370	PASS
175	174	5	9	7.5	6792	PASS
176	174	95	101	97.0	87637	PASS
177	176	5	9	7.5	6584	PASS