

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060221\
 Data File : VY004923.D
 Acq On : 02 Jun 2021 13:56
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY060221

Quant Time: Jun 03 04:46:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.454	0.459	-1.1	106	0.00
3 P	Chloromethane	0.620	0.636	-2.6	106	0.00
4 C	Vinyl Chloride	0.596	0.620	-4.0#	105	0.00
5 T	Bromomethane	0.338	0.362	-7.1	112	0.00
6 T	Chloroethane	0.357	0.375	-5.0	106	0.00
7 T	Trichlorofluoromethane	0.844	0.898	-6.4	109	0.00
8 T	Diethyl Ether	0.328	0.349	-6.4	108	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.515	0.552	-7.2	110	0.00
10 T	Methyl Iodide	0.491	0.552	-12.4	110	0.00
11 T	Tert butyl alcohol	0.061	0.061	0.0	104	0.00
12 CM	1,1-Dichloroethene	0.514	0.543	-5.6#	107	0.00
13 T	Acrolein	0.062	0.060	3.2	100	0.00
14 T	Allyl chloride	0.937	1.001	-6.8	107	0.00
15 T	Acrylonitrile	0.154	0.169	-9.7	104	0.00
16 T	Acetone	0.127	0.134	-5.5	106	0.00
17 T	Carbon Disulfide	1.580	1.674	-5.9	106	0.00
18 T	Methyl Acetate	0.408	0.426	-4.4	103	0.00
19 T	Methyl tert-butyl Ether	1.323	1.430	-8.1	105	0.00
20 T	Methylene Chloride	0.741	0.674	9.0	109	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.608	-5.4	106	0.00
22 T	Diisopropyl ether	1.720	1.849	-7.5	106	0.00
23 T	Vinyl Acetate	1.251	1.367	-9.3	104	0.00
24 P	1,1-Dichloroethane	1.035	1.102	-6.5	107	0.00
25 T	2-Butanone	0.214	0.233	-8.9	103	0.00
26 T	2,2-Dichloropropane	0.875	0.872	0.3	102	0.00
27 T	cis-1,2-Dichloroethene	0.632	0.634	-0.3	102	0.00
28 T	Bromochloromethane	0.462	0.474	-2.6	99	0.00
29 T	Tetrahydrofuran	0.136	0.149	-9.6	103	0.00
30 C	Chloroform	0.982	1.038	-5.7#	105	0.00
31 T	Cyclohexane	0.931	0.958	-2.9	110	0.00
32 T	1,1,1-Trichloroethane	0.861	0.902	-4.8	104	0.00
33 S	1,2-Dichloroethane-d4	0.529	0.532	-0.6	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.269	0.274	-1.9	97	0.00
36 T	1,1-Dichloropropene	0.489	0.519	-6.1	107	0.00
37 T	Ethyl Acetate	0.322	0.324	-0.6	101	0.00
38 T	Carbon Tetrachloride	0.426	0.463	-8.7	106	0.00
39 T	Methylcyclohexane	0.541	0.601	-11.1	113	0.00
40 TM	Benzene	1.400	1.461	-4.4	104	0.00
41 T	Methacrylonitrile	0.177	0.163	7.9	87	0.00
42 TM	1,2-Dichloroethane	0.404	0.428	-5.9	105	0.00
43 T	Isopropyl Acetate	0.578	0.620	-7.3	103	0.00
44 TM	Trichloroethene	0.366	0.384	-4.9	106	0.00
45 C	1,2-Dichloropropane	0.366	0.390	-6.6#	108	0.00
46 T	Dibromomethane	0.196	0.208	-6.1	106	0.00
47 T	Bromodichloromethane	0.459	0.493	-7.4	109	0.00
48 T	Methyl methacrylate	0.263	0.282	-7.2	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060221\
 Data File : VY004923.D
 Acq On : 02 Jun 2021 13:56
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY060221

Quant Time: Jun 03 04:46:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	107	0.00
50 S	Toluene-d8	1.117	1.131	-1.3	98	0.00
51 T	4-Methyl-2-Pentanone	0.303	0.324	-6.9	102	0.00
52 CM	Toluene	0.847	0.886	-4.6#	104	0.00
53 T	t-1,3-Dichloropropene	0.509	0.531	-4.3	103	0.00
54 T	cis-1,3-Dichloropropene	0.578	0.602	-4.2	104	0.00
55 T	1,1,2-Trichloroethane	0.272	0.281	-3.3	103	0.00
56 T	Ethyl methacrylate	0.389	0.420	-8.0	101	0.00
57 T	1,3-Dichloropropane	0.484	0.506	-4.5	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.197	0.203	-3.0	101	0.00
59 T	2-Hexanone	0.205	0.224	-9.3	102	0.00
60 T	Dibromochloromethane	0.309	0.327	-5.8	104	0.00
61 T	1,2-Dibromoethane	0.261	0.272	-4.2	103	0.00
62 S	4-Bromofluorobenzene	0.384	0.392	-2.1	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.418	0.443	-6.0	106	0.00
65 PM	Chlorobenzene	0.991	1.042	-5.1	106	0.00
66 T	1,1,1,2-Tetrachloroethane	0.353	0.375	-6.2	106	0.00
67 C	Ethyl Benzene	1.831	1.950	-6.5#	106	0.00
68 T	m/p-Xylenes	0.686	0.737	-7.4	106	0.00
69 T	o-Xylene	0.639	0.694	-8.6	108	0.00
70 T	Styrene	1.064	1.173	-10.2	107	0.00
71 P	Bromoform	0.208	0.224	-7.7	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.777	4.166	-10.3	110	0.00
74 T	N-amyl acetate	1.219	1.337	-9.7	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.758	0.801	-5.7	103	0.00
76 T	1,2,3-Trichloropropane	0.548	0.557	-1.6	99	0.00
77 T	Bromobenzene	0.844	0.911	-7.9	105	0.00
78 T	n-propylbenzene	4.553	5.043	-10.8	110	0.00
79 T	2-Chlorotoluene	2.538	2.767	-9.0	107	0.00
80 T	1,3,5-Trimethylbenzene	3.116	3.395	-9.0	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.298	0.320	-7.4	103	0.00
82 T	4-Chlorotoluene	2.593	2.812	-8.4	107	0.00
83 T	tert-Butylbenzene	2.636	2.891	-9.7	107	0.00
84 T	1,2,4-Trimethylbenzene	3.097	3.418	-10.4	108	0.00
85 T	sec-Butylbenzene	3.837	4.216	-9.9	110	0.00
86 T	p-Isopropyltoluene	3.292	3.686	-12.0	111	0.00
87 T	1,3-Dichlorobenzene	1.606	1.744	-8.6	108	0.00
88 T	1,4-Dichlorobenzene	1.602	1.715	-7.1	108	0.00
89 T	n-Butylbenzene	3.166	3.544	-11.9	112	0.00
90 T	Hexachloroethane	0.489	0.530	-8.4	110	0.00
91 T	1,2-Dichlorobenzene	1.419	1.557	-9.7	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.135	-7.1	105	0.00
93 T	1,2,4-Trichlorobenzene	0.891	0.976	-9.5	107	0.00
94 T	Hexachlorobutadiene	0.470	0.526	-11.9	107	0.00
95 T	Naphthalene	1.960	2.219	-13.2	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060221\
Data File : VY004923.D
Acq On : 02 Jun 2021 13:56
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY060221

Quant Time: Jun 03 04:46:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.785	0.886	-12.9	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6