

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060222\
 Data File : VY009129.D
 Acq On : 02 Jun 2022 15:51
 Operator : KP/MD
 Sample : N3139-03
 Misc : 3.98g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Jun 03 03:12:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 20 23:50:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	111017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	199609	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	192283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	84022	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	74435	65.147	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.300%
35) Dibromofluoromethane	7.716	113	64573	52.022	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.040%
50) Toluene-d8	10.179	98	247736	54.101	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.200%
62) 4-Bromofluorobenzene	12.477	95	72539	39.270	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060222\
Data File : VY009129.D
Acq On : 02 Jun 2022 15:51
Operator : KP/MD
Sample : N3139-03
Misc : 3.98g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Jun 03 03:12:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
Quant Title : SW846 8260
QLast Update : Fri May 20 23:50:52 2022
Response via : Initial Calibration

