



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060222\
 Data File : VY009118.D
 Acq On : 02 Jun 2022 10:08
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052022S.M
 Title : SW846 8260
 Last Update : Fri May 20 23:50:52 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.9	7679	PASS
75	95	30	60	58.3	18720	PASS
95	95	100	100	100.0	32083	PASS
96	95	5	9	7.0	2247	PASS
173	174	0.00	2	1.6	470	PASS
174	95	50	100	91.6	29379	PASS
175	174	5	9	8.1	2377	PASS
176	174	95	101	96.5	28365	PASS
177	176	5	9	6.6	1880	PASS