

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
 Data File : VY013984.D
 Acq On : 02 Jun 2023 12:59
 Operator : KP/MD
 Sample : 03002-04
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB04

Quant Time: Jun 03 04:59:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	1999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	3448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	3666	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	732	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	2107	79.623	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	159.240%
35) Dibromofluoromethane	7.716	113	1453	66.309	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	132.620%
50) Toluene-d8	10.179	98	4135	49.472	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.940%
62) 4-Bromofluorobenzene	12.477	95	1059	37.164	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	74.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
Data File : VY013984.D
Acq On : 02 Jun 2023 12:59
Operator : KP/MD
Sample : 03002-04
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB04

Quant Time: Jun 03 04:59:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

