

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
 Data File : VY013987.D
 Acq On : 02 Jun 2023 14:10
 Operator : KP/MD
 Sample : 03006-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-3-060123

Quant Time: Jun 05 09:42:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	112689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	185519	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	187007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	86906	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	79236	53.117	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.240%
35) Dibromofluoromethane	7.710	113	65267	55.358	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.720%
50) Toluene-d8	10.173	98	230644	51.286	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.580%
62) 4-Bromofluorobenzene	12.477	95	75190	49.042	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
Data File : VY013987.D
Acq On : 02 Jun 2023 14:10
Operator : KP/MD
Sample : 03006-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-3-060123

Quant Time: Jun 05 09:42:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

