

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
 Data File : VY013992.D
 Acq On : 02 Jun 2023 16:07
 Operator : KP/MD
 Sample : 03008-01
 Misc : 5.66g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-18-0A

Quant Time: Jun 05 09:44:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	159825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	274151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	287125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	145385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	121259	57.314	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	114.620%	
35) Dibromofluoromethane	7.710	113	95447	54.783	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	109.560%	
50) Toluene-d8	10.173	98	344111	51.779	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	103.560%	
62) 4-Bromofluorobenzene	12.477	95	127323	56.197	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	112.400%	
Target Compounds						Qvalue
16) Acetone	3.948	43	39822	55.014	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
Data File : VY013992.D
Acq On : 02 Jun 2023 16:07
Operator : KP/MD
Sample : 03008-01
Misc : 5.66g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-18-0A

Quant Time: Jun 05 09:44:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

