

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
 Data File : VY013993.D
 Acq On : 02 Jun 2023 16:31
 Operator : KP/MD
 Sample : 03008-04
 Misc : 5.54g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-19-0A

Quant Time: Jun 05 09:44:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	159440	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	278935	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	287030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	141442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	119595	56.664	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.320%	
35) Dibromofluoromethane	7.709	113	95296	53.759	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.520%	
50) Toluene-d8	10.172	98	351389	51.968	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 103.940%	
62) 4-Bromofluorobenzene	12.477	95	123328	53.500	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 107.000%	
Target Compounds						
16) Acetone	3.960	43	12680	17.560	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
Data File : VY013993.D
Acq On : 02 Jun 2023 16:31
Operator : KP/MD
Sample : 03008-04
Misc : 5.54g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-19-0A

Quant Time: Jun 05 09:44:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

