

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
 Data File : VY013995.D
 Acq On : 02 Jun 2023 17:18
 Operator : KP/MD
 Sample : 03016-03
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUP04

Quant Time: Jun 05 09:45:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 02 07:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	4720	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.679	114	7574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	7501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2446	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3860	61.778	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.560%
35) Dibromofluoromethane	7.703	113	2813	58.441	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.880%
50) Toluene-d8	10.179	98	8882	48.376	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.760%
62) 4-Bromofluorobenzene	12.471	95	2596	41.474	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060223\
Data File : VY013995.D
Acq On : 02 Jun 2023 17:18
Operator : KP/MD
Sample : 03016-03
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUP04

Quant Time: Jun 05 09:45:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060123S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 02 07:00:14 2023
Response via : Initial Calibration

