

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
 Data File : VY022503.D
 Acq On : 02 Jun 2025 17:34
 Operator : SY/MD
 Sample : Q2152-01RE
 Misc : 5.37g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-02-05292025

Quant Time: Jun 03 04:08:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.707	168	5540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	9515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	6126	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	891	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	3259	52.597	ug/l	-0.01
Spiked Amount 50.000	Range 50 - 163		Recovery = 105.200%			
35) Dibromofluoromethane	7.640	113	2037	35.867	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 71.740%			
50) Toluene-d8	10.109	98	10756	46.871	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 93.740%			
62) 4-Bromofluorobenzene	12.408	95	1359	19.876	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 39.760%			
Target Compounds						
					Qvalue	
3) Chloromethane	2.068	50	1071	7.309	ug/l	89
4) Vinyl Chloride	2.178	62	402	2.365	ug/l	# 85
5) Bromomethane	2.629	94	793	4.867	ug/l	# 60
10) Methyl Iodide	4.025	142	866	13.167	ug/l	# 42
11) Tert butyl alcohol	4.598	59	64	14.462	ug/l	# 73
13) Acrolein	3.671	56	20	3.626	ug/l	# 14
15) Acrylonitrile	5.116	53	31	2.252	ug/l	# 23
16) Acetone	3.867	43	121	9.616	ug/l	# 45
18) Methyl Acetate	4.391	43	176	4.723	ug/l	# 51
25) 2-Butanone	6.903	43	25	1.377	ug/l	# 46
29) Tetrahydrofuran	7.262	42	54	4.613	ug/l	# 31
36) 1,1-Dichloropropene	7.701	75	518	5.620	ug/l	# 42
37) Ethyl Acetate	6.994	43	106	2.579	ug/l	# 69
38) Carbon Tetrachloride	7.720	117	524	5.542	ug/l	# 13
41) Methacrylonitrile	7.256	41	74	3.082	ug/l	# 6
48) Methyl methacrylate	9.256	41	160	3.840	ug/l	# 11
49) 1,4-Dioxane	9.463	88	23	53.634	ug/l	# 11
51) 4-Methyl-2-Pentanone	10.000	43	68	1.536	ug/l	# 37
61) 1,2-Dibromoethane	11.048	107	51	1.212	ug/l	# 1
74) N-amyl acetate	11.993	43	51	2.763	ug/l	# 43
76) 1,2,3-Trichloropropane	12.645	75	43	4.218	ug/l	# 100
78) n-propylbenzene	12.597	91	147	1.645	ug/l	90
79) 2-Chlorotoluene	12.694	91	103	2.175	ug/l	# 41
81) trans-1,4-Dichloro-2-b...	12.414	75	831	196.875	ug/l	# 10
82) 4-Chlorotoluene	12.786	91	183	3.746	ug/l	# 41
84) 1,2,4-Trimethylbenzene	13.182	105	73	1.271	ug/l	# 31
91) 1,2-Dichlorobenzene	13.657	146	54	2.034	ug/l	# 47
93) 1,2,4-Trichlorobenzene	14.919	180	51	3.521	ug/l	# 6
95) Naphthalene	15.151	128	212	7.628	ug/l	# 69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
Data File : VY022503.D
Acq On : 02 Jun 2025 17:34
Operator : SY/MD
Sample : Q2152-01RE
Misc : 5.37g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025

Quant Time: Jun 03 04:08:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 99
Delta R.T. -0.006 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion:168 Resp: 5540
Ion Ratio Lower Upper
168 100
99 54.3 44.3 66.5



#3
Chloromethane
Concen: 7.309 ug/l
RT: 2.068 min Scan# 57
Delta R.T. 0.000 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 50 Resp: 1071
Ion Ratio Lower Upper
50 100
52 26.5 26.0 39.0









#4
 Vinyl Chloride
 Concen: 2.365 ug/l
 RT: 2.178 min Scan# 71
 Delta R.T. -0.024 min
 Lab File: VY022503.D
 Acq: 02 Jun 2025 17:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-02-05292025


Tgt Ion: 62 Resp: 402
 Ion Ratio Lower Upper
 62 100
 64 23.6 25.3 37.9#



#5
 Bromomethane
 Concen: 4.867 ug/l
 RT: 2.629 min Scan# 149
 Delta R.T. 0.030 min
 Lab File: VY022503.D
 Acq: 02 Jun 2025 17:34

Tgt Ion: 94 Resp: 793
 Ion Ratio Lower Upper
 94 100
 96 53.8 74.0 111.0#









#10
Methyl Iodide
Concen: 13.167 ug/l
RT: 4.025 min Scan# 31
Delta R.T. 0.018 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion:142 Resp: 866
Ion Ratio Lower Upper
142 100
127 0.0 37.1 55.7#
141 6.7 11.4 17.2#



#11
Tert butyl alcohol
Concen: 14.462 ug/l
RT: 4.598 min Scan# 472
Delta R.T. -0.287 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 59 Resp: 64
Ion Ratio Lower Upper
59 100
57 0.0 8.0 12.0#









#13
Acrolein
Concen: 3.626 ug/l
RT: 3.671 min Scan# 318
Delta R.T. 0.012 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion: 56 Resp: 20
Ion Ratio Lower Upper
56 100
55 0.0 57.0 85.6#



#15
Acrylonitrile
Concen: 2.252 ug/l
RT: 5.116 min Scan# 557
Delta R.T. 0.049 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 53 Resp: 31
Ion Ratio Lower Upper
53 100
52 0.0 65.7 98.5#
51 64.5 30.1 45.1#









#16
Acetone
Concen: 9.616 ug/l
RT: 3.867 min Scan# 31
Delta R.T. -0.012 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion: 43 Resp: 121
Ion Ratio Lower Upper
43 100
58 0.0 24.0 36.0#



#18
Methyl Acetate
Concen: 4.723 ug/l
RT: 4.391 min Scan# 438
Delta R.T. -0.006 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 43 Resp: 176
Ion Ratio Lower Upper
43 100
74 0.0 19.8 29.6#









#25
2-Butanone
Concen: 1.377 ug/l
RT: 6.903 min Scan# 81
Delta R.T. 0.000 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion: 43 Resp: 25
Ion Ratio Lower Upper
43 100
72 0.0 22.9 34.3#



#29
Tetrahydrofuran
Concen: 4.613 ug/l
RT: 7.262 min Scan# 909
Delta R.T. -0.006 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 42 Resp: 54
Ion Ratio Lower Upper
42 100
72 0.0 37.4 56.2#
71 0.0 34.2 51.2#





#33

1,2-Dichloroethane-d4

Concen: 52.597 ug/l

RT: 8.055 min Scan# 1103

Delta R.T. -0.012 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34

Instrument :

MSVOA_Y

ClientSampleId :





OK-02-05292025

Tgt Ion: 65 Resp: 3259

Ion Ratio Lower Upper

65 100

67 56.3 0.0 103.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34




Tgt Ion: 114 Resp: 9515


Ion Ratio Lower Upper

114 100

63 16.9 0.0 40.8

88 16.9 0.0 27.8









#35
Dibromofluoromethane
Concen: 35.867 ug/l
RT: 7.640 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion:113 Resp: 2037
Ion Ratio Lower Upper
113 100
111 114.6 81.1 121.7
192 16.0 14.2 21.2



#36
1,1-Dichloropropene
Concen: 5.620 ug/l
RT: 7.701 min Scan# 981
Delta R.T. -0.140 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 75 Resp: 518
Ion Ratio Lower Upper
75 100
110 0.0 16.9 50.7#
77 0.0 24.8 37.2#





#37

Ethyl Acetate

Concen: 2.579 ug/l

RT: 6.994 min Scan# 80

Delta R.T. 0.000 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34

Instrument :

MSVOA_Y

ClientSampleId :

OK-02-05292025





Tgt Ion: 43 Resp: 106

Ion Ratio Lower Upper

43 100

61 0.0 11.5 17.3#

70 0.0 7.2 10.8#



#38

Carbon Tetrachloride

Concen: 5.542 ug/l

RT: 7.720 min Scan# 984

Delta R.T. -0.104 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34




Tgt Ion: 117 Resp: 524


Ion Ratio Lower Upper

117 100

119 0.0 75.0 112.6#


121 0.0 25.0 37.4#








#41
Methacrylonitrile
Concen: 3.082 ug/l
RT: 7.256 min Scan# 904
Delta R.T. 0.024 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025







Tgt Ion:	41	Resp:	74
Ion	Ratio	Lower	Upper
41	100		
39	147.3	46.6	69.8#
67	0.0	63.3	94.9#
52	0.0	27.5	41.3#



#48
Methyl methacrylate
Concen: 3.840 ug/l
RT: 9.256 min Scan# 1236
Delta R.T. 0.030 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion:	41	Resp:	160
Ion	Ratio	Lower	Upper
41	100		
69	0.0	67.8	101.8#
39	0.0	55.0	82.4#





#49

1,4-Dioxane

Concen: 53.634 ug/l

RT: 9.463 min Scan# 11

Delta R.T. 0.219 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34

Instrument :

MSVOA_Y

ClientSampleId :

OK-02-05292025





Tgt Ion: 88 Resp: 23

Ion Ratio Lower Upper

88 100

43 87.0 25.4 38.2#

58 0.0 54.9 82.3#



#50

Toluene-d8

Concen: 46.871 ug/l

RT: 10.109 min Scan# 1376

Delta R.T. 0.000 min

Lab File: VY022503.D




Acq: 02 Jun 2025 17:34


Tgt Ion: 98 Resp: 10756

Ion Ratio Lower Upper

98 100

100 60.7 51.4 77.0





#51

4-Methyl-2-Pentanone

Concen: 1.536 ug/l

RT: 10.000 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34

Instrument :

MSVOA_Y

ClientSampleId :





OK-02-05292025

Tgt Ion: 43 Resp: 68

Ion Ratio Lower Upper

43 100

58 0.0 30.4 45.6#



#61

1,2-Dibromoethane

Concen: 1.212 ug/l

RT: 11.048 min Scan# 1530

Delta R.T. 0.030 min

Lab File: VY022503.D




Acq: 02 Jun 2025 17:34


Tgt Ion: 107 Resp: 51

Ion Ratio Lower Upper

107 100

109 0.0 77.7 116.5#









#62
4-Bromofluorobenzene
Concen: 19.876 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion:	95	Resp:	1359
Ion	Ratio	Lower	Upper
95	100		
174	81.6	0.0	170.0
176	103.0	0.0	166.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 1591
Delta R.T. 0.000 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion:	117	Resp:	6126
Ion	Ratio	Lower	Upper
117	100		
82	72.0	44.6	66.8#
119	38.5	25.4	38.0#





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.353 min Scan# 1907

Delta R.T. 0.006 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34

Instrument :

MSVOA_Y

ClientSampleId :

OK-02-05292025





Tgt Ion:152 Resp: 891

Ion Ratio Lower Upper

152 100

115 58.5 28.9 86.7

150 153.2 0.0 349.6



#74

N-ethyl acetate

Concen: 2.763 ug/l

RT: 11.993 min Scan# 1685

Delta R.T. -0.079 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34

Tgt Ion: 43 Resp: 51




Ion Ratio Lower Upper


43 100

70 0.0 34.2 51.2#

55 0.0 18.2 27.4#

61 0.0 19.0 28.6#









#76
1,2,3-Trichloropropane
Concen: 4.218 ug/l
RT: 12.645 min Scan# 1777
Delta R.T. 0.091 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion: 75 Resp: 43
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0



#78
n-propylbenzene
Concen: 1.645 ug/l
RT: 12.597 min Scan# 1784
Delta R.T. 0.000 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 91 Resp: 147
Ion Ratio Lower Upper
91 100
120 17.0 10.8 32.4









#79
2-Chlorotoluene
Concen: 2.175 ug/l
RT: 12.694 min Scan# 1798
Delta R.T. 0.012 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion: 91 Resp: 103
Ion Ratio Lower Upper
91 100
126 0.0 17.0 51.0#



#81
trans-1,4-Dichloro-2-butene
Concen: 196.875 ug/l
RT: 12.414 min Scan# 1754
Delta R.T. 0.110 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 75 Resp: 831
Ion Ratio Lower Upper
75 100
53 0.0 78.5 117.7#
89 0.0 35.7 53.5#









#82
4-Chlorotoluene
Concen: 3.746 ug/l
RT: 12.786 min Scan# 1814
Delta R.T. 0.006 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34




Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025


Tgt Ion: 91 Resp: 183
Ion Ratio Lower Upper
91 100
126 0.0 17.0 50.9#



#84
1,2,4-Trimethylbenzene
Concen: 1.271 ug/l
RT: 13.182 min Scan# 1880
Delta R.T. 0.140 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Tgt Ion: 105 Resp: 73
Ion Ratio Lower Upper
105 100
120 0.0 22.7 68.1#





#91

1,2-Dichlorobenzene

Concen: 2.034 ug/l

RT: 13.657 min Scan# 1958

Delta R.T. 0.000 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34





Instrument :

MSVOA_Y

ClientSampleId :

OK-02-05292025

Tgt Ion:146	Resp:	54
Ion Ratio	Lower	Upper
146	100	
111	46.3	20.2 60.6
148	0.0	31.4 94.0#



#93

1,2,4-Trichlorobenzene

Concen: 3.521 ug/l




RT: 14.919 min Scan# 2165


Delta R.T. 0.000 min

Lab File: VY022503.D

Acq: 02 Jun 2025 17:34

Tgt Ion:180	Resp:	51
Ion Ratio	Lower	Upper
180	100	
182	0.0	48.4 145.3#
145	74.5	15.5 46.5#





#95
Naphthalene
Concen: 7.628 ug/l
RT: 15.151 min Scan# 21
Delta R.T. 0.006 min
Lab File: VY022503.D
Acq: 02 Jun 2025 17:34

Instrument :
MSVOA_Y
ClientSampleId :
OK-02-05292025

Tgt Ion:128 Resp: 212
Ion Ratio Lower Upper
128 100
127 0.0 10.4 15.6#
129 0.0 8.9 13.3#

