

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
 Data File : VY022505.D
 Acq On : 02 Jun 2025 18:21
 Operator : SY/MD
 Sample : Q2147-02
 Misc : 6.00g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Jun 03 05:32:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	1372	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	2319	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	647	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	163	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	1029	67.057	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.120%
35) Dibromofluoromethane	7.634	113	628	45.370	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.740%
50) Toluene-d8	10.115	98	1857	33.203	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	66.400%
62) 4-Bromofluorobenzene	12.420	95	302	18.123	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	36.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
Data File : VY022505.D
Acq On : 02 Jun 2025 18:21
Operator : SY/MD
Sample : Q2147-02
Misc : 6.00g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Jun 03 05:32:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration

