

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
 Data File : VY022506.D
 Acq On : 02 Jun 2025 18:44
 Operator : SY/MD
 Sample : Q2150-02
 Misc : 6.52g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-42

Quant Time: Jun 03 04:09:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	279093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	514728	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	411552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	142944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	170529	54.630	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 109.260%		
35) Dibromofluoromethane	7.634	113	159229	51.827	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 103.660%		
50) Toluene-d8	10.109	98	621468	50.062	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 100.120%		
62) 4-Bromofluorobenzene	12.408	95	154197	41.689	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 83.380%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	4.878	59	4623	20.736	ug/l	96
16) Acetone	3.885	43	14916	23.529	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
Data File : VY022506.D
Acq On : 02 Jun 2025 18:44
Operator : SY/MD
Sample : Q2150-02
Misc : 6.52g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-42

Quant Time: Jun 03 04:09:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration

