

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY060225\  
 Data File : VY022492.D  
 Acq On : 02 Jun 2025 12:09  
 Operator : SY/MD  
 Sample : VSTDIC010  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :

Quant Time: Jun 03 02:54:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y060225S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 03 02:50:07 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
 Supervised By :Mahesh Dadoda 06/03/2025

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|------------------------------|--------|-------|----------|----------|--------|----------|
| -----                        |        |       |          |          |        |          |
| Internal Standards           |        |       |          |          |        |          |
| 1) Pentafluorobenzene        | 7.713  | 168   | 211235   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.622  | 114   | 366386   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.420 | 117   | 302825   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.352 | 152   | 137338   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds  |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.067  | 65    | 24246    | 10.549   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 50 - 163 | Recovery | =      | 21.100%# |
| 35) Dibromofluoromethane     | 7.646  | 113   | 20743    | 9.617    | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 54 - 147 | Recovery | =      | 19.240%# |
| 50) Toluene-d8               | 10.109 | 98    | 86561    | 9.930    | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 58 - 134 | Recovery | =      | 19.860%# |
| 62) 4-Bromofluorobenzene     | 12.408 | 95    | 24870    | 9.409    | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 30 - 143 | Recovery | =      | 18.820%# |
| Target Compounds             |        |       |          |          |        |          |
|                              |        |       |          |          |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.867  | 85    | 22911    | 12.412   | ug/l   | 99       |
| 3) Chloromethane             | 2.068  | 50    | 67191    | 12.339   | ug/l   | 99       |
| 4) Vinyl Chloride            | 2.202  | 62    | 76656    | 12.027   | ug/l   | 97       |
| 5) Bromomethane              | 2.598  | 94    | 73289    | 12.157   | ug/l   | 99       |
| 6) Chloroethane              | 2.739  | 64    | 53992    | 12.374   | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 3.056  | 101   | 63044    | 11.642   | ug/l   | 97       |
| 8) Diethyl Ether             | 3.458  | 74    | 14144    | 11.826   | ug/l   | 100      |
| 9) 1,1,2-Trichlorotrifluo... | 3.824  | 101   | 25997    | 11.445   | ug/l   | 99       |
| 10) Methyl Iodide            | 4.007  | 142   | 25590    | 10.505   | ug/l   | 99       |
| 11) Tert butyl alcohol       | 4.891  | 59    | 9732     | 59.275   | ug/l # | 73       |
| 12) 1,1-Dichloroethene       | 3.799  | 96    | 24370    | 11.240   | ug/l   | 99       |
| 13) Acrolein                 | 3.671  | 56    | 9650     | 48.151   | ug/l   | 94       |
| 14) Allyl chloride           | 4.391  | 41    | 38482    | 11.389   | ug/l   | 95       |
| 15) Acrylonitrile            | 5.073  | 53    | 29150    | 57.299   | ug/l   | 99       |
| 16) Acetone                  | 3.885  | 43    | 25066    | 55.166   | ug/l   | 98       |
| 17) Carbon Disulfide         | 4.110  | 76    | 79008    | 11.413   | ug/l   | 99       |
| 18) Methyl Acetate           | 4.397  | 43    | 17228    | 12.603   | ug/l   | 97       |
| 19) Methyl tert-butyl Ether  | 5.128  | 73    | 68336    | 11.296   | ug/l   | 95       |
| 20) Methylene Chloride       | 4.622  | 84    | 29569    | 11.655   | ug/l   | 96       |
| 21) trans-1,2-Dichloroethene | 5.128  | 96    | 27493    | 11.350   | ug/l   | 88       |
| 22) Diisopropyl ether        | 6.031  | 45    | 83814    | 11.156   | ug/l   | 96       |
| 23) Vinyl Acetate            | 5.970  | 43    | 241364   | 54.152   | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 5.927  | 63    | 50391    | 11.346   | ug/l   | 96       |
| 25) 2-Butanone               | 6.909  | 43    | 36756    | 54.840   | ug/l   | 95       |
| 26) 2,2-Dichloropropane      | 6.890  | 77    | 43605    | 11.045   | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 6.902  | 96    | 30873    | 11.028   | ug/l   | 99       |
| 28) Bromochloromethane       | 7.256  | 49    | 20132    | 10.491   | ug/l   | 99       |
| 29) Tetrahydrofuran          | 7.274  | 42    | 25001    | 57.822   | ug/l   | 97       |
| 30) Chloroform               | 7.427  | 83    | 49990    | 11.391   | ug/l   | 89       |
| 31) Cyclohexane              | 7.707  | 56    | 51050    | 11.422   | ug/l # | 95       |
| 32) 1,1,1-Trichloroethane    | 7.628  | 97    | 43454    | 11.080   | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 7.841  | 75    | 37532    | 10.765   | ug/l   | 99       |
| 37) Ethyl Acetate            | 6.994  | 43    | 15914    | 10.371   | ug/l # | 93       |
| 38) Carbon Tetrachloride     | 7.829  | 117   | 37794    | 10.515   | ug/l   | 100      |
| 39) Methylcyclohexane        | 9.115  | 83    | 49413    | 10.491   | ug/l   | 97       |
| 40) Benzene                  | 8.085  | 78    | 108934   | 10.600   | ug/l   | 96       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY060225\  
 Data File : VY022492.D  
 Acq On : 02 Jun 2025 12:09  
 Operator : SY/MD  
 Sample : VSTDIC010  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :

Quant Time: Jun 03 02:54:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y060225S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 03 02:50:07 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
 Supervised By :Mahesh Dadoda 06/03/2025

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.238  | 41   | 9043m    | 9.151   | ug/l   | 96       |
| 42) 1,2-Dichloroethane        | 8.164  | 62   | 31102    | 11.125  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 8.201  | 43   | 36411    | 10.859  | ug/l   | 97       |
| 44) Trichloroethene           | 8.865  | 130  | 28431    | 11.325  | ug/l   | 91       |
| 45) 1,2-Dichloropropane       | 9.146  | 63   | 26665    | 10.884  | ug/l   | 97       |
| 46) Dibromomethane            | 9.237  | 93   | 14601m   | 10.637  | ug/l   | 96       |
| 47) Bromodichloromethane      | 9.426  | 83   | 38475    | 11.048  | ug/l   | 98       |
| 48) Methyl methacrylate       | 9.225  | 41   | 16122    | 10.501  | ug/l   | 97       |
| 49) 1,4-Dioxane               | 9.243  | 88   | 3535     | 215.881 | ug/l   | 100      |
| 51) 4-Methyl-2-Pentanone      | 9.999  | 43   | 88634    | 53.259  | ug/l   | 99       |
| 52) Toluene                   | 10.176 | 92   | 68265    | 10.565  | ug/l   | 95       |
| 53) t-1,3-Dichloropropene     | 10.396 | 75   | 34307    | 10.649  | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 9.859  | 75   | 40229    | 10.649  | ug/l   | 94       |
| 55) 1,1,2-Trichloroethane     | 10.579 | 97   | 18848    | 10.904  | ug/l   | 97       |
| 56) Ethyl methacrylate        | 10.444 | 69   | 26171    | 10.426  | ug/l   | 99       |
| 57) 1,3-Dichloropropane       | 10.719 | 76   | 33836    | 10.991  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 9.713  | 63   | 52396    | 46.819  | ug/l   | 100      |
| 59) 2-Hexanone                | 10.761 | 43   | 57738    | 52.045  | ug/l   | 99       |
| 60) Dibromochloromethane      | 10.914 | 129  | 23423    | 10.694  | ug/l   | 96       |
| 61) 1,2-Dibromoethane         | 11.018 | 107  | 17302    | 10.847  | ug/l   | 97       |
| 64) Tetrachloroethene         | 10.652 | 164  | 28680    | 11.144  | ug/l   | 99       |
| 65) Chlorobenzene             | 11.444 | 112  | 70845    | 10.789  | ug/l   | 96       |
| 66) 1,1,1,2-Tetrachloroethane | 11.517 | 131  | 23209    | 10.427  | ug/l   | 100      |
| 67) Ethyl Benzene             | 11.524 | 91   | 126556   | 10.350  | ug/l   | 99       |
| 68) m/p-Xylenes               | 11.633 | 106  | 94535    | 20.427  | ug/l   | 99       |
| 69) o-Xylene                  | 11.956 | 106  | 45374    | 10.476  | ug/l   | 98       |
| 70) Styrene                   | 11.975 | 104  | 72373    | 10.092  | ug/l   | 99       |
| 71) Bromoform                 | 12.133 | 173  | 12637    | 10.804  | ug/l # | 99       |
| 73) Isopropylbenzene          | 12.255 | 105  | 118344   | 10.695  | ug/l   | 99       |
| 74) N-amyl acetate            | 12.072 | 43   | 29609    | 10.581  | ug/l   | 96       |
| 75) 1,1,2,2-Tetrachloroethane | 12.511 | 83   | 20264    | 11.276  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 12.560 | 75   | 16751m   | 9.239   | ug/l   | 98       |
| 77) Bromobenzene              | 12.536 | 156  | 25392    | 10.827  | ug/l   | 100      |
| 78) n-propylbenzene           | 12.597 | 91   | 143414   | 10.648  | ug/l   | 98       |
| 79) 2-Chlorotoluene           | 12.682 | 91   | 75325    | 10.591  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 12.737 | 105  | 91121    | 10.532  | ug/l   | 96       |
| 81) trans-1,4-Dichloro-2-b... | 12.304 | 75   | 7009     | 11.084  | ug/l   | 100      |
| 82) 4-Chlorotoluene           | 12.779 | 91   | 78689    | 10.676  | ug/l   | 97       |
| 83) tert-Butylbenzene         | 12.999 | 119  | 81823    | 10.367  | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 13.048 | 105  | 90848    | 10.495  | ug/l   | 100      |
| 85) sec-Butylbenzene          | 13.176 | 105  | 127776   | 10.678  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 13.291 | 119  | 99940    | 10.158  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 13.291 | 146  | 48971    | 10.437  | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene       | 13.371 | 146  | 49834    | 10.900  | ug/l   | 99       |
| 89) n-Butylbenzene            | 13.621 | 91   | 97544    | 10.413  | ug/l   | 96       |
| 90) Hexachloroethane          | 13.877 | 117  | 20466    | 10.631  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 13.663 | 146  | 43460    | 10.781  | ug/l   | 95       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.273 | 75   | 3014     | 11.095  | ug/l   | 98       |
| 93) 1,2,4-Trichlorobenzene    | 14.919 | 180  | 23710    | 10.760  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 15.023 | 225  | 13179    | 10.951  | ug/l   | 99       |
| 95) Naphthalene               | 15.145 | 128  | 44061    | 10.434  | ug/l   | 96       |
| 96) 1,2,3-Trichlorobenzene    | 15.328 | 180  | 19814    | 10.581  | ug/l   |          |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY060225\  
Data File : VY022492.D  
Acq On : 02 Jun 2025 12:09  
Operator : SY/MD  
Sample : VSTDIC010  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jun 03 02:54:00 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y060225S.M  
Quant Title : SW846 8260  
QLast Update : Tue Jun 03 02:50:07 2025  
Response via : Initial Calibration

Instrument :  
MSVOA\_Y  
ClientSampleId :

Manual Integrations  
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025

| Compound                                                                         | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------------------------------------------------------------|------|------|----------|------|-------|----------|
| -----                                                                            |      |      |          |      |       |          |
| ( # ) = qualifier out of range ( m ) = manual integration ( + ) = signals summed |      |      |          |      |       |          |

06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.713 min Scan# 91  
Delta R.T. 0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:168 Resp: 211235  
Ion Ratio Lower Upper  
168 100  
99 58.9 44.3 66.5


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda

06/03/2025



#2  
Dichlorodifluoromethane  
Concen: 12.412 ug/l  
RT: 1.867 min Scan# 24  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 85 Resp: 22911  
Ion Ratio Lower Upper  
85 100  
87 33.7 16.7 50.0





#3  
Chloromethane  
Concen: 12.339 ug/l  
RT: 2.068 min Scan# 51  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 50 Resp: 6719  
Ion Ratio Lower Upper  
50 100  
52 33.1 26.0 39.0


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#4  
Vinyl Chloride  
Concen: 12.027 ug/l  
RT: 2.202 min Scan# 79  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion: 62 Resp: 76656  
Ion Ratio Lower Upper  
62 100  
64 30.1 25.3 37.9





#5  
Bromomethane  
Concen: 12.157 ug/l  
RT: 2.598 min Scan# 144  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 94 Resp: 73289  
Ion Ratio Lower Upper  
94 100  
96 91.1 74.0 111.0


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#6  
Chloroethane  
Concen: 12.374 ug/l  
RT: 2.739 min Scan# 167  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion: 64 Resp: 53992  
Ion Ratio Lower Upper  
64 100  
66 30.3 23.8 35.6





#7  
Trichlorofluoromethane  
Concen: 11.642 ug/l  
RT: 3.056 min Scan# 219  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:101 Resp: 63044  
Ion Ratio Lower Upper  
101 100  
103 62.5 52.2 78.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025



#8  
Diethyl Ether  
Concen: 11.826 ug/l  
RT: 3.458 min Scan# 285  
Delta R.T. -0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Tgt Ion: 74 Resp: 14144  
Ion Ratio Lower Upper  
74 100  
45 91.9 45.9 137.6







#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 11.445 ug/l  
RT: 3.824 min Scan# 344  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Instrument :  
MSVOA\_Y  
ClientSampleId :

Tgt Ion:101 Resp: 2599  
Ion Ratio Lower Upper  
101 100  
85 48.6 39.3 58.9  
151 71.3 57.8 86.8



Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#10  
Methyl Iodide  
Concen: 10.505 ug/l  
RT: 4.007 min Scan# 375  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion:142 Resp: 25590  
Ion Ratio Lower Upper  
142 100  
127 47.3 37.1 55.7  
141 14.6 11.4 17.2





#11  
Tert butyl alcohol  
Concen: 59.275 ug/l  
RT: 4.891 min Scan# 519  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 59 Resp: 973  
Ion Ratio Lower Upper  
59 100  
57 0.0 8.0 12.0


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#12  
1,1-Dichloroethene  
Concen: 11.240 ug/l  
RT: 3.799 min Scan# 341  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion: 96 Resp: 24370  
Ion Ratio Lower Upper  
96 100  
61 167.3 134.6 201.8  
98 61.7 50.2 75.4





#13  
Acrolein  
Concen: 48.151 ug/l  
RT: 3.671 min Scan# 318  
Delta R.T. 0.012 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 56 Resp: 9650  
Ion Ratio Lower Upper  
56 100  
55 76.0 57.0 85.6


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#14  
Allyl chloride  
Concen: 11.389 ug/l  
RT: 4.391 min Scan# 438  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 41 Resp: 38482  
Ion Ratio Lower Upper  
41 100  
39 80.5 69.1 103.7  
76 39.7 33.5 50.3





#15  
Acrylonitrile  
Concen: 57.299 ug/l  
RT: 5.073 min Scan# 515  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 53 Resp: 29150  
Ion Ratio Lower Upper  
53 100  
52 82.7 65.7 98.5  
51 38.3 30.1 45.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#16  
Acetone  
Concen: 55.166 ug/l  
RT: 3.885 min Scan# 355  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion: 43 Resp: 25066  
Ion Ratio Lower Upper  
43 100  
58 31.1 24.0 36.0





#17  
Carbon Disulfide  
Concen: 11.413 ug/l  
RT: 4.110 min Scan# 392  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 76 Resp: 79008  
Ion Ratio Lower Upper  
76 100  
78 8.8 7.4 11.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#18  
Methyl Acetate  
Concen: 12.603 ug/l  
RT: 4.397 min Scan# 439  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion: 43 Resp: 17228  
Ion Ratio Lower Upper  
43 100  
74 26.4 19.8 29.6





#19  
Methyl tert-butyl Ether  
Concen: 11.296 ug/l  
RT: 5.128 min Scan# 51  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 73 Resp: 68336  
Ion Ratio Lower Upper  
73 100  
57 23.9 17.4 26.0


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#20  
Methylene Chloride  
Concen: 11.655 ug/l  
RT: 4.622 min Scan# 476  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 84 Resp: 29569  
Ion Ratio Lower Upper  
84 100  
49 123.4 101.9 152.9  
51 37.5 29.8 44.8  
86 57.3 51.3 76.9





#21  
trans-1,2-Dichloroethene  
Concen: 11.350 ug/l  
RT: 5.128 min Scan# 51  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 96 Resp: 2749  
Ion Ratio Lower Upper  
96 100  
61 123.0 112.6 169.0  
98 57.1 50.6 75.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#22  
Diisopropyl ether  
Concen: 11.156 ug/l  
RT: 6.031 min Scan# 707  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 45 Resp: 83814  
Ion Ratio Lower Upper  
45 100  
43 64.5 48.3 72.5  
87 28.5 23.0 34.6  
59 12.7 9.4 14.0





#23  
 Vinyl Acetate  
 Concen: 54.152 ug/l  
 RT: 5.970 min Scan# 697  
 Delta R.T. -0.000 min  
 Lab File: VY022492.D  
 Acq: 02 Jun 2025 12:09



Instrument :  
 MSVOA\_Y  
 ClientSampleId :



Tgt Ion: 43 Resp: 241364  
 Ion Ratio Lower Upper  
 43 100  
 86 10.9 8.6 12.8


Manual Integrations  
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
 Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
 Supervised By :Mahesh  
 Dadoda



06/03/2025



#24  
 1,1-Dichloroethane  
 Concen: 11.346 ug/l  
 RT: 5.927 min Scan# 690  
 Delta R.T. 0.006 min  
 Lab File: VY022492.D  
 Acq: 02 Jun 2025 12:09




Tgt Ion: 63 Resp: 50391  
 Ion Ratio Lower Upper  
 63 100  
 98 5.0 3.4 10.1  
 100 3.0 2.1 6.3





#25  
2-Butanone  
Concen: 54.840 ug/l  
RT: 6.909 min Scan# 851  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 43 Resp: 36750  
Ion Ratio Lower Upper  
43 100  
72 31.4 22.9 34.3


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda

06/03/2025



#26  
2,2-Dichloropropane  
Concen: 11.045 ug/l  
RT: 6.890 min Scan# 848  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 77 Resp: 43605  
Ion Ratio Lower Upper  
77 100  
97 22.4 11.4 34.2





#27  
cis-1,2-Dichloroethene  
Concen: 11.028 ug/l  
RT: 6.902 min Scan# 81  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 96 Resp: 3087  
Ion Ratio Lower Upper  
96 100  
61 145.8 0.0 290.0  
98 65.3 0.0 128.4


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#28  
Bromochloromethane  
Concen: 10.491 ug/l  
RT: 7.256 min Scan# 908  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 49 Resp: 20132  
Ion Ratio Lower Upper  
49 100  
129 1.5 0.0 3.6  
130 70.8 55.7 83.5





#29  
Tetrahydrofuran  
Concen: 57.822 ug/l  
RT: 7.274 min Scan# 911  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 42 Resp: 2500  
Ion Ratio Lower Upper  
42 100  
72 43.5 37.4 56.2  
71 42.5 34.2 51.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#30  
Chloroform  
Concen: 11.391 ug/l  
RT: 7.427 min Scan# 936  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 83 Resp: 49990  
Ion Ratio Lower Upper  
83 100  
85 59.0 54.2 81.4





#31  
Cyclohexane  
Concen: 11.422 ug/l  
RT: 7.707 min Scan# 982  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 56 Resp: 51050  
Ion Ratio Lower Upper  
56 100  
69 40.9 27.0 40.6  
84 92.7 72.6 109.0


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#32  
1,1,1-Trichloroethane  
Concen: 11.080 ug/l  
RT: 7.628 min Scan# 969  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Tgt Ion: 97 Resp: 43454  
Ion Ratio Lower Upper  
97 100  
99 65.1 50.9 76.3  
61 50.6 38.5 57.7





#33

1,2-Dichloroethane-d4

Concen: 10.549 ug/l

RT: 8.067 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

Instrument :

MSVOA\_Y

ClientSampleId :

Tgt Ion: 65 Resp: 24240

Ion Ratio Lower Upper

65 100



67 51.1 0.0 103.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025



Supervised By :Mahesh Dadoda 06/03/2025



06/03/2025

Supervised By :Mahesh  
Dadoda

06/03/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.622 min Scan# 1132

Delta R.T. 0.006 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09




Tgt Ion:114 Resp: 366386

Ion Ratio Lower Upper

114 100

63 20.6 0.0 40.8


88 13.7 0.0 27.8





#35  
Dibromofluoromethane  
Concen: 9.617 ug/l  
RT: 7.646 min Scan# 91  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 113 Resp: 2074  
Ion Ratio Lower Upper  
113 100  
111 104.9 81.1 121.7  
192 18.7 14.2 21.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#36  
1,1-Dichloropropene  
Concen: 10.765 ug/l  
RT: 7.841 min Scan# 1004  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 75 Resp: 37532  
Ion Ratio Lower Upper  
75 100  
110 33.0 16.9 50.7  
77 31.5 24.8 37.2





#37  
Ethyl Acetate  
Concen: 10.371 ug/l  
RT: 6.994 min Scan# 80  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 43 Resp: 15914  
Ion Ratio Lower Upper  
43 100  
61 18.4 11.5 17.3  
70 9.8 7.2 10.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#38  
Carbon Tetrachloride  
Concen: 10.515 ug/l  
RT: 7.829 min Scan# 1002  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion:117 Resp: 37794  
Ion Ratio Lower Upper  
117 100  
119 94.2 75.0 112.6  
121 30.9 25.0 37.4





#39  
Methylcyclohexane  
Concen: 10.491 ug/l  
RT: 9.115 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 83 Resp: 4941  
Ion Ratio Lower Upper  
83 100  
55 75.8 58.1 87.1  
98 47.5 39.0 58.6


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#40  
Benzene  
Concen: 10.600 ug/l  
RT: 8.085 min Scan# 1044  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion: 78 Resp: 108934  
Ion Ratio Lower Upper  
78 100  
77 26.2 19.3 28.9





#41  
Methacrylonitrile  
Concen: 9.151 ug/l m  
RT: 7.238 min Scan# 904  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 41 Resp: 904  
Ion Ratio Lower Upper  
41 100  
39 83.2 46.6 69.8  
67 0.0 63.3 94.9  
52 0.0 27.5 41.3#


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#42  
1,2-Dichloroethane  
Concen: 11.125 ug/l  
RT: 8.164 min Scan# 1057  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 62 Resp: 31102  
Ion Ratio Lower Upper  
62 100  
98 9.9 0.0 19.2





#43  
Isopropyl Acetate  
Concen: 10.859 ug/l  
RT: 8.201 min Scan# 1063  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 43 Resp: 3641  
Ion Ratio Lower Upper  
43 100  
61 17.7 15.6 23.4  
87 13.4 10.2 15.4


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#44  
Trichloroethene  
Concen: 11.325 ug/l  
RT: 8.865 min Scan# 1172  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Tgt Ion:130 Resp: 28431  
Ion Ratio Lower Upper  
130 100  
95 92.0 0.0 203.0





#45

1,2-Dichloropropane

Concen: 10.884 ug/l

RT: 9.146 min Scan# 1218

Delta R.T. -0.000 min


Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

Instrument :

MSVOA\_Y

ClientSampleId :



Tgt Ion: 63 Resp: 2666

Ion Ratio Lower Upper

63 100



65 28.2 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025


Supervised By :Mahesh Dadoda 06/03/2025



06/03/2025

Supervised By :Mahesh  
Dadoda

06/03/2025



#46

Dibromomethane

Concen: 10.637 ug/l m

RT: 9.237 min Scan# 1233

Delta R.T. 0.006 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09




Tgt Ion: 93 Resp: 14601


Ion Ratio Lower Upper

93 100

95 0.0 65.6 98.4#


174 0.0 70.4 105.6#





#47  
Bromodichloromethane  
Concen: 11.048 ug/l  
RT: 9.426 min Scan# 1264  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 83 Resp: 3847  
Ion Ratio Lower Upper  
83 100  
85 57.8 49.1 73.7  
127 8.4 6.2 9.4


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#48  
Methyl methacrylate  
Concen: 10.501 ug/l  
RT: 9.225 min Scan# 1231  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion: 41 Resp: 16122  
Ion Ratio Lower Upper  
41 100  
69 86.6 67.8 101.8  
39 70.2 55.0 82.4





#49  
1,4-Dioxane  
Concen: 215.881 ug/l  
RT: 9.243 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 88 Resp: 353  
Ion Ratio Lower Upper  
88 100  
43 36.1 25.4 38.2  
58 69.8 54.9 82.3


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#50  
Toluene-d8  
Concen: 9.930 ug/l  
RT: 10.109 min Scan# 1376  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 98 Resp: 86561  
Ion Ratio Lower Upper  
98 100  
100 63.2 51.4 77.0





#51  
4-Methyl-2-Pentanone  
Concen: 53.259 ug/l  
RT: 9.999 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 43 Resp: 88634  
Ion Ratio Lower Upper  
43 100  
58 38.3 30.4 45.6


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#52  
Toluene  
Concen: 10.565 ug/l  
RT: 10.176 min Scan# 1387  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 92 Resp: 68265  
Ion Ratio Lower Upper  
92 100  
91 172.7 139.0 208.4





#53  
t-1,3-Dichloropropene  
Concen: 10.649 ug/l  
RT: 10.396 min Scan# 1423  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 75 Resp: 3430  
Ion Ratio Lower Upper  
75 100  
77 30.3 26.5 39.7


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#54  
cis-1,3-Dichloropropene  
Concen: 10.649 ug/l  
RT: 9.859 min Scan# 1335  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 75 Resp: 40229  
Ion Ratio Lower Upper  
75 100  
77 28.7 25.3 37.9  
39 52.8 42.7 64.1





#55  
1,1,2-Trichloroethane  
Concen: 10.904 ug/l  
RT: 10.579 min Scan# 1452  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 97 Resp: 1884  
Ion Ratio Lower Upper  
97 100  
83 90.5 68.1 102.1  
85 55.3 48.2 72.2  
99 67.2 49.5 74.3


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda

06/03/2025



#56  
Ethyl methacrylate  
Concen: 10.426 ug/l  
RT: 10.444 min Scan# 1431  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Tgt Ion: 69 Resp: 26171  
Ion Ratio Lower Upper  
69 100  
41 67.3 52.2 78.2  
39 47.9 36.7 55.1





#57

1,3-Dichloropropane

Concen: 10.991 ug/l

RT: 10.719 min Scan# 1476

Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

Instrument :



MSVOA\_Y

ClientSampleId :

Tgt Ion: 76 Resp: 33830  
 Ion Ratio Lower Upper  
 76 100  
 78 31.8 25.8 38.8



Manual Integrations  
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
 Supervised By :Mahesh Dadoda 06/03/2025



06/03/2025  
 Supervised By :Mahesh  
 Dadoda

06/03/2025



#58

2-Chloroethyl Vinyl ether

Concen: 46.819 ug/l




RT: 9.713 min Scan# 1311


Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09


Tgt Ion: 63 Resp: 52396  
 Ion Ratio Lower Upper  
 63 100  
 106 27.7 21.6 32.4





#59  
2-Hexanone  
Concen: 52.045 ug/l  
RT: 10.761 min Scan# 1483  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 43 Resp: 57738  
Ion Ratio Lower Upper  
43 100  
58 54.3 27.3 81.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#60  
Dibromochloromethane  
Concen: 10.694 ug/l  
RT: 10.914 min Scan# 1508  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion:129 Resp: 23423  
Ion Ratio Lower Upper  
129 100  
127 79.3 39.1 117.2





#61  
1,2-Dibromoethane  
Concen: 10.847 ug/l  
RT: 11.018 min Scan# 1525  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:107 Resp: 1730  
Ion Ratio Lower Upper  
107 100  
109 93.0 77.7 116.5


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#62  
4-Bromofluorobenzene  
Concen: 9.409 ug/l  
RT: 12.408 min Scan# 1753  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Tgt Ion: 95 Resp: 24870  
Ion Ratio Lower Upper  
95 100  
174 83.4 0.0 170.0  
176 82.2 0.0 166.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.420 min Scan# 1591

Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

Instrument :



MSVOA\_Y

ClientSampleId :

Tgt Ion:117 Resp: 302825  
 Ion Ratio Lower Upper  
 117 100  
 82 56.6 44.6 66.8  
 119 31.7 25.4 38.0



Manual Integrations  
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
 Supervised By :Mahesh Dadoda 06/03/2025



06/03/2025  
 Supervised By :Mahesh  
 Dadoda

06/03/2025



#64

Tetrachloroethene

Concen: 11.144 ug/l




RT: 10.652 min Scan# 1465


Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

Tgt Ion:164 Resp: 28680  
 Ion Ratio Lower Upper  
 164 100  
 166 127.3 104.0 156.0  
 129 96.5 74.5 111.7  
 131 87.6 72.3 108.5







#65  
Chlorobenzene  
Concen: 10.789 ug/l  
RT: 11.444 min Scan# 1595  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Instrument :  
MSVOA\_Y  
ClientSampleId :

Tgt Ion:112 Resp: 70849  
Ion Ratio Lower Upper  
112 100  
114 31.9 25.0 37.6



Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda

06/03/2025



#66  
1,1,1,2-Tetrachloroethane  
Concen: 10.427 ug/l  
RT: 11.517 min Scan# 1607  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Tgt Ion:131 Resp: 23209  
Ion Ratio Lower Upper  
131 100  
133 100.5 47.3 141.8  
119 66.5 33.1 99.3







#67  
Ethyl Benzene  
Concen: 10.350 ug/l  
RT: 11.524 min Scan# 1608  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Instrument :  
MSVOA\_Y  
ClientSampleId :

Tgt Ion: 91 Resp: 126550  
Ion Ratio Lower Upper  
91 100  
106 30.3 24.2 36.4



Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#68  
m/p-Xylenes  
Concen: 20.427 ug/l  
RT: 11.633 min Scan# 1626  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion:106 Resp: 94535  
Ion Ratio Lower Upper  
106 100  
91 200.7 159.4 239.0





#69  
o-Xylene  
Concen: 10.476 ug/l  
RT: 11.956 min Scan# 1679  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:106 Resp: 45374  
Ion Ratio Lower Upper  
106 100  
91 209.3 105.8 317.3


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#70  
Styrene  
Concen: 10.092 ug/l  
RT: 11.975 min Scan# 1682  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion:104 Resp: 72373  
Ion Ratio Lower Upper  
104 100  
78 49.8 39.5 59.3  
103 56.9 44.1 66.1





#71  
Bromoform  
Concen: 10.804 ug/l  
RT: 12.133 min Scan# 1708  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:173 Resp: 1263  
Ion Ratio Lower Upper  
173 100  
175 48.6 23.8 71.5  
254 0.0 0.0 0.0


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.352 min Scan# 1908  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion:152 Resp: 137338  
Ion Ratio Lower Upper  
152 100  
115 58.4 28.9 86.7  
150 162.7 0.0 349.6





#73  
Isopropylbenzene  
Concen: 10.695 ug/l  
RT: 12.255 min Scan# 11834  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:105 Resp: 11834  
Ion Ratio Lower Upper  
105 100  
120 26.2 13.0 38.9


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#74  
N-ethyl acetate  
Concen: 10.581 ug/l  
RT: 12.072 min Scan# 1698  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Tgt Ion: 43 Resp: 29609  
Ion Ratio Lower Upper  
43 100  
70 42.8 34.2 51.2  
55 23.8 18.2 27.4  
61 22.9 19.0 28.6







#75  
1,1,2,2-Tetrachloroethane  
Concen: 11.276 ug/l  
RT: 12.511 min Scan# 1769  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Instrument :  
MSVOA\_Y  
ClientSampleId :

Tgt Ion: 83 Resp: 20264  
Ion Ratio Lower Upper  
83 100  
131 9.0 4.9 14.7  
85 62.2 32.6 97.8



Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#76  
1,2,3-Trichloropropane  
Concen: 9.239 ug/l m  
RT: 12.560 min Scan# 1778  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 75 Resp: 16751  
Ion Ratio Lower Upper  
75 100  
77 0.0 0.0 0.0





#77  
Bromobenzene  
Concen: 10.827 ug/l  
RT: 12.536 min Scan# 1774  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:156 Resp: 25392  
Ion Ratio Lower Upper  
156 100  
77 191.7 94.0 282.1  
158 97.6 48.2 144.6


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#78  
n-propylbenzene  
Concen: 10.648 ug/l  
RT: 12.597 min Scan# 1784  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 91 Resp: 143414  
Ion Ratio Lower Upper  
91 100  
120 21.8 10.8 32.4





#79  
2-Chlorotoluene  
Concen: 10.591 ug/l  
RT: 12.682 min Scan# 1798  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 91 Resp: 7532  
Ion Ratio Lower Upper  
91 100  
126 35.0 17.0 51.0


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#80  
1,3,5-Trimethylbenzene  
Concen: 10.532 ug/l  
RT: 12.737 min Scan# 1807  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion:105 Resp: 91121  
Ion Ratio Lower Upper  
105 100  
120 49.2 24.1 72.4





#81  
trans-1,4-Dichloro-2-butene  
Concen: 11.084 ug/l  
RT: 12.304 min Scan# 1736  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 75 Resp: 7009  
Ion Ratio Lower Upper  
75 100  
53 94.7 78.5 117.7  
89 47.6 35.7 53.5


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#82  
4-Chlorotoluene  
Concen: 10.676 ug/l  
RT: 12.779 min Scan# 1814  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion: 91 Resp: 78689  
Ion Ratio Lower Upper  
91 100  
126 34.2 17.0 50.9





#83  
tert-Butylbenzene  
Concen: 10.367 ug/l  
RT: 12.999 min Scan# 1850  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:119 Resp: 8182  
Ion Ratio Lower Upper  
119 100  
91 66.3 32.1 96.3  
134 27.2 13.0 38.9


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#84  
1,2,4-Trimethylbenzene  
Concen: 10.495 ug/l  
RT: 13.048 min Scan# 1858  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Tgt Ion:105 Resp: 90848  
Ion Ratio Lower Upper  
105 100  
120 46.5 22.7 68.1





#85

sec-Butylbenzene

Concen: 10.678 ug/l

RT: 13.176 min Scan# 1879

Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

Instrument :

MSVOA\_Y

ClientSampleId :

Tgt Ion:105 Resp: 127770

Ion Ratio Lower Upper

105 100



134 19.3 9.7 28.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025



Supervised By :Mahesh Dadoda 06/03/2025



06/03/2025

Supervised By :Mahesh  
Dadoda

06/03/2025



#86

p-Isopropyltoluene

Concen: 10.158 ug/l

RT: 13.291 min Scan# 1898

Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09




Tgt Ion:119 Resp: 99940


Ion Ratio Lower Upper

119 100

134 25.2 13.1 39.1

91 23.5 11.6 34.8





#87

1,3-Dichlorobenzene

Concen: 10.437 ug/l

RT: 13.291 min Scan# 1898

Delta R.T. -0.000 min


Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

Instrument :

MSVOA\_Y



ClientSampleId :



|           |       |       |      |
|-----------|-------|-------|------|
| Tgt Ion:  | 146   | Resp: | 4897 |
| Ion Ratio | Lower | Upper |      |
| 146       | 100   |       |      |
| 111       | 40.6  | 20.1  | 60.2 |
| 148       | 64.6  | 32.1  | 96.3 |


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025



06/03/2025  
Supervised By :Mahesh  
Dadoda

06/03/2025



#88

1,4-Dichlorobenzene

Concen: 10.900 ug/l




RT: 13.371 min Scan# 1911


Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09


|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 146   | Resp: | 49834 |
| Ion Ratio | Lower | Upper |       |
| 146       | 100   |       |       |
| 111       | 39.2  | 19.4  | 58.2  |
| 148       | 60.6  | 31.6  | 94.8  |





#89  
n-Butylbenzene  
Concen: 10.413 ug/l  
RT: 13.621 min Scan# 1952  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion: 91 Resp: 97544  
Ion Ratio Lower Upper  
91 100  
92 51.0 25.9 77.8  
134 24.6 12.0 36.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#90  
Hexachloroethane  
Concen: 10.631 ug/l  
RT: 13.877 min Scan# 1994  
Delta R.T. -0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09


Tgt Ion:117 Resp: 20466  
Ion Ratio Lower Upper  
117 100  
201 64.9 34.0 102.0





#91  
1,2-Dichlorobenzene  
Concen: 10.781 ug/l  
RT: 13.663 min Scan# 1958  
Delta R.T. 0.006 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:146 Resp: 43460  
Ion Ratio Lower Upper  
146 100  
111 40.0 20.2 60.6  
148 64.0 31.4 94.0


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025




06/03/2025  
Supervised By :Mahesh  
Dadoda



06/03/2025




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 11.095 ug/l  
RT: 14.273 min Scan# 2059  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09




Tgt Ion: 75 Resp: 3014  
Ion Ratio Lower Upper  
75 100  
155 86.2 39.9 119.6  
157 111.3 54.0 162.0





#93  
1,2,4-Trichlorobenzene  
Concen: 10.760 ug/l  
RT: 14.919 min Scan# 2165  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09



Instrument :  
MSVOA\_Y  
ClientSampleId :



Tgt Ion:180 Resp: 23710  
Ion Ratio Lower Upper  
180 100  
182 94.9 48.4 145.3  
145 29.5 15.5 46.5


Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025






06/03/2025  
Supervised By :Mahesh  
Dadoda


06/03/2025



#94  
Hexachlorobutadiene  
Concen: 10.951 ug/l  
RT: 15.023 min Scan# 2182  
Delta R.T. -0.000 min  
Lab File: VY022492.D  
Acq: 02 Jun 2025 12:09

Tgt Ion:225 Resp: 13179  
Ion Ratio Lower Upper  
225 100  
223 63.6 32.5 97.4  
227 64.7 32.3 96.9





#95

Naphthalene

Concen: 10.434 ug/l

RT: 15.145 min Scan# 21

Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

Instrument :



MSVOA\_Y

ClientSampleId :

|             |       |           |
|-------------|-------|-----------|
| Tgt Ion:128 | Resp: | 4406      |
| Ion Ratio   | Lower | Upper     |
| 128         | 100   |           |
| 127         | 13.4  | 10.4 15.6 |
| 129         | 10.9  | 8.9 13.3  |



Manual Integrations  
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025  
Supervised By :Mahesh Dadoda 06/03/2025



06/03/2025  
Supervised By :Mahesh  
Dadoda

06/03/2025



#96

1,2,3-Trichlorobenzene

Concen: 10.581 ug/l

RT: 15.328 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY022492.D

Acq: 02 Jun 2025 12:09

|             |       |            |
|-------------|-------|------------|
| Tgt Ion:180 | Resp: | 19814      |
| Ion Ratio   | Lower | Upper      |
| 180         | 100   |            |
| 182         | 97.1  | 47.1 141.3 |
| 145         | 33.5  | 15.2 45.6  |

