

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060321\
 Data File : VY004951.D
 Acq On : 03 Jun 2021 20:51
 Operator : SY/MD
 Sample : M2582-01
 Misc : 8.03G/5ML/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-1-VNJ-201

Quant Time: Jun 04 01:20:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 03 04:43:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	199055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	367794	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	351961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	160395	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	138800	65.958	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.920%
35) Dibromofluoromethane	7.728	113	114671	57.950	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.900%
50) Toluene-d8	10.185	98	459136	55.869	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.740%
62) 4-Bromofluorobenzene	12.483	95	164216	58.102	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.200%
Target Compounds						
95) Naphthalene	15.232	128	10818	1.721	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060321\
Data File : VY004951.D
Acq On : 03 Jun 2021 20:51
Operator : SY/MD
Sample : M2582-01
Misc : 8.03G/5ML/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-1-VNJ-201

Quant Time: Jun 04 01:20:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060221S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 03 04:43:33 2021
Response via : Initial Calibration

